

HIN/MED/2019-2020/100602
15th April 2019

CALIBRATION CERTIFICATE

This is to certify that the hematology analyser **MICROS EMI CRP** bearing serial no: **45901004** installed at **ASAVLEE DR. APARNA JAIRAM PATHOLOGY & LAB, MALAD EAST MUMBAI** was calibrated on 15th April 2019.

Calibrator : ABX MINOCAL
Lot No. : CX433
Expiry Date : 05th May 2019

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 14th April 2020.



Shrish Dixit

(Head – Products & Marketing)

For Horiba India Pvt. Ltd.