

METHOD (1)

	Units	Level 1 - 26441		Level 2 - 26442	
		Mean	Range	Mean	Range
PHENYTOIN					
Emit 2000	µg/mL	§		§	
PHOSPHORUS					
Catachem ChemWell-T (4)	mg/dL	4.10	3.30 – 4.90	7.00	6.00 – 8.00
Phosphomolybdate-UV (3)	mg/dL	3.54	3.10 – 3.98	7.36	6.70 – 8.02
Teco Diagnostics TC-Matrix (Molybdenum UV)	mg/dL	▲		▲	
POTASSIUM					
Flame Photometry	mEq/L	3.83	3.43 – 4.23	6.09	5.02 – 7.15
ISE Indirect (3)	mEq/L	3.85	3.61 – 4.09	6.15	5.81 – 6.50
Teco Diagnostics TC-Matrix (Turbidimetric)	mEq/L	▲		▲	
PROTEIN SERUM (TOTAL)					
Riuret. no serum blank. end point (3)	g/dl	6.61	5.92 – 7.30	4.35	3.83 – 4.87

READING SHOWN IN THE MACHINE FOR THE IQC PHOSPHOROUS SAMPLE 29/02/2020

COLLATED REPORT BY DATE.					
DATE:	29/02/20				
000010	29/02/20	AMY	1164	IU/l	H
000011	29/02/20	PHOS	3.29	mg/dl	
000012	29/02/20	CA-A	9.75	mg/dl	