

# TSH, uU/ml

**Method:** Beckman Access/LXI725 hyper TSH 3rd gen.  
**Instrument:** Beckman Access Series  
**Reagent:** Beckman Coulter

**RIQAS TDPA:** 18.8%      **Biological Variation:** 23.7%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	2.840	uU/ml	25	1.143	6.1	0.02	0.13	12.99	10	148.49
2	0.950	uU/ml	30	1.011	6.0	0.01	0.12	-0.52	100	-6.00
3	1.218	uU/ml	29	1.246	8.6	0.02	0.14	-0.19	120	-2.21
4	0.173	uU/ml	29	0.179	6.4	0.00	0.02	-0.28	120	-3.20
5	23.034	uU/ml	28	24.377	4.9	0.28	2.79	-0.48	103	-5.51
6	1.164	uU/ml	24	1.192	5.1	0.02	0.14	-0.21	120	-2.38
7	0.952	uU/ml	26	0.905	7.0	0.02	0.10	0.46	105	5.25
8	1.385	uU/ml	28	1.344	7.1	0.02	0.15	0.26	120	3.03
9	8.590	uU/ml	28	1.060	6.4	0.02	0.12	62.18	10	710.30
10	1.037	uU/ml	26	1.011	6.4	0.02	0.12	0.23	120	2.61
11	1.293	uU/ml	28	1.306	7.8	0.02	0.15	-0.09	120	-1.03
12	1.205	uU/ml	29	1.103	6.4	0.02	0.13	0.81	81	9.30

	Cycle 15	Cycle 16
Cycle Average SDI	-0.02	6.26
Cycle Average TS	101	94
Cycle Average %DEV	-0.25	71.55
Cycle Average Absolute SDI	0.50	6.56
Cycle Average Absolute %DEV	5.75	74.94

