

Calibration Certificate

Following voltages have been checked & adjusted on Blood Cell Counter SWELAB ALFA (SI.No.23115) at Day & Night Diagnostics (Sanjeevan Medical Research Centre) 4869/24, Ansari Road Daryaganj, New Delhi-110002.

Power Supply: Voltages 24 V are OK.

Checked the Gain and Offset of Analog Board and found OK.

All instrument thru' the Service Program and found OK.

Checked the Pumps & Valves, Shear Valve and found OK.

Parameters background check is also under acceptable range.

Calibrated the instrument using Boule Calibrator Lot No-21901-34.

Calibration Report

Parameter	Target	Range	Actual Result
RBC($10^{10}/L$)	3.91	0.10	3.96
MCV(fL)	82.6	2.0	82.5
Plt($10^9/L$)	236	15.0	231
MPV(fL)	11.0	1.5	11.6
WBC($10^9/L$)	8.2	0.3	8.3
Hgb(g/dl)	11.5	0.2	11.5

Next Calibration will be performed on 15/04/2020

Regards
For, AAMX BIOMEDICALS

Date: 14/04/2019




Man Singh
Head Customer Service
931021234