

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

ULR - CC279519000014090F		Page 1 of 1
Certificate Number : PMM/MP/6565/3		12.06.2019
Calibrated For	M/s. Sanjeevan Medical Research Centre (P) Ltd 24, Ansari Road, Daryaganj, New Delhi - 02	14.06.2019 13.06.2020 15.06.2019
		UUC Received on Calibrated on Valid Upto Cert. Issue Date

Description of Sample			
Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	100 to 1000 µl	Condition of UUC	Physically Ok
Resolution	10 µl	Calibration Performed at	Laboratory
Make & Model No.	Vertex	Calibration Procedure No.	CP/M/MV/04
Serial No.	190340935	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	_____	Environmental Condition	Temperature = 23 ± 2° C Humidity = 50 ± 10% RH
Location	Lab		

Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL	PMM/WB/1139	21.05.2020

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Mechanical Calibration

CALIBRATION RESULTS				
Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	100	100.00	0.00	100.00
2	500	501.79	-1.79	501.82
3	1000	1004.33	-4.33	1004.38

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 5.78 µl

Priyank Sharma

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Jai Prakash

Approved By : Jai Prakash
 Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.