



# Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307  
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com  
Website : www.pmmpl.in, www.parashar.info  
(NABL ACCREDITED LABORATORY)

## Calibration Certificate

Certificate Number : CAL/WB/S/503816		Date of Request	Page 1 of 1
Calibrated For	M/s. Sanjeevan Medical Research Centre (P) Ltd. 24, Ansari Road, Daryaganj, New Delhi - 02	Calibrated on	11.06.2019
		Valid Upto	12.06.2019
		Cert. Issue Date	11.06.2020
			15.06.2019

Description of Sample			
Nomenclature	Water Bath (Thermostat)	Accuracy/Class	Not Mentioned
Range	0 to 110 ° C	Condition of UUC	Physically Ok
Resolution	5° C	Calibration Performed at	At Site
Make & Model No.	-----	Calibration Procedure No.	CP/T - 03
Serial No.	-----	Reference IS Standard	IS:7358 - 1984
Party ID Mark	SMRC / Lab / NWB / 01	Environmental Condition	Temperature = 25 ± 5° C
Location	Lab		Humidity = 40 to 70% RH

Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03 PMM/CTR/01	Tempsens Udaipur	CC284019000003882F	12.05.2020

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Thermal Calibration			
CALIBRATION RESULTS			
S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	40	39.35	0.65
2	50	48.74	1.26
3	80	78.16	1.84
4	90	87.59	2.41
5	100	97.02	2.98

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor k = 2) = ± 3 ° C

*Priyank*

Calibrated By : Priyank Sharma  
Designation : Calibration Engineer



*Moh. Iqbal*

Approved By : Moh. Iqbal  
Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
  2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
  3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.