

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)

Calibration Certificate

Page 1 of 1

Certificate Number : CAL/DPG/S/503811		Date of Request	11.06.2019			
Calibrated For	M/s. Sanjeevan Medical Research Centre (P) Ltd. 24, Ansari Road, Daryaganj, New Delhi - 02	Calibrated on	12.06.2019			
		Valid Upto	11.06.2020			
		Cert. Issue Date	15.06.2019			
Description of Sample						
Nomenclature	Laminar Air Flow	Accuracy/Class	Not Mentioned			
Range	0 to 50 mmWc	Condition of UUC	Physically Ok			
Resolution	1.25 mmWc	Calibration Performed at	At Site			
Make & Model No.	Vaiometra	Calibration Procedure No.	CP/MP/S/01			
Serial No.	-----	Reference IS Standard	IS: DKD - R6 - 1			
Party ID Mark	SMRC / Lab / LAF / 01	Environmental Condition	Temperature = 23 ± 5° C			
Location	Lab		Humidity = 40 to 70%RH			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Pressure Calibrator	Polltech	19214	Adcon, Gurgaon	ATCL/18-19/ME-492	21.12.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Mechanical Calibration


CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in mmWc)	Actual Observed Value in Std. (in mmWc)	Error (in mmWc)
1	5	5	0
2	10	10	0
3	15	15	0
4	20	20	0
5	25	25	0
6	30	29	1


C - Unit Under Calibration

- Standard Instrument

ertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 2 mmWc

Calibrated By: 
 Priyank Sharma
 Calibration Engineer



Approved By: 
 Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.