



Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)

Calibration Certificate

Page 1 of 1
 11.06.2019
 12.06.2019
11.06.2020
 15.06.2019

Certificate Number : CAL/CF/S/503809		Date of Request
Calibrated For	M/s. Sanjeevan Medical Research Centre (P) Ltd. 24, Ansari Road, Daryaganj, New Delhi - 02	Calibrated on
		Valid Upto
		Cert. Issue Date

Description of Sample

Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned
Range	Refer to Result	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	At Site
Make & Model No.	Remi	Calibration Procedure No.	CP/M/MV/S/03
Serial No.	FELC - 6694	Reference IS Standard	-----
Party ID Mark	SMRC / Lab / Centri/02	Environmental Condition	Temperature = 23 ± 5° C
Location	Lab		Humidity = 40 to 70%RH

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Lutron	PMM/TM/01	PMMPL, Noida	PMM/TM/1075	11.09.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Mechanical Calibration

CALIBRATION RESULTS

Sr.No.	Step	Actual Observed Value in Std. (in rpm)	± Uncertainty in Measurement (in rpm)
1	1	1521	45
2	2	2138	45
3	3	2625	45
4	4	2946	45
5	5	3490	45

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Sharma

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
 Designation : Tech. Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the
 2. The calibration