



Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)

Calibration Certificate

Certificate Number : CAL/IN/S/503814		Date of Request	Page 1 of 1
Calibrated For	M/s Sanjeevan Medical Research Centre (P) Ltd 24, Ansari Road, Daryaganj, New Delhi - 02	Calibrated on	11.06.2019
		Valid Upto	12.06.2019
		Cert. Issue Date	11.06.2020
			15.06.2019

Description of Sample			
Nomenclature	B.O.D Incubator (Dig. Temp. Controller With Sensor)	Accuracy/Class	Not Mentioned
Range	0 to 99.9 ° C	Condition of UUC	Physically Ok
Resolution	0.1° C	Calibration Performed at	At Site
Make & Model No.	Vaiometra	Calibration Procedure No.	CP/T/S/04
Serial No.	F0043640914	Reference IS Standard	IS:2806 - 1992
Party ID Mark	SMRC / Lab / BOD INC / 01	Environmental Condition	Temperature = 25 ± 5° C
Location	Lab		Humidity = 40 to 70% RH

Master Equipment / Standard Used						
Sr.N.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03 PMM/CTR/01	Tempsens Udaipur	CC284019000003882F	12.05.2020

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Thermal Calibration

CALIBRATION RESULTS

S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	35.1	35.03	0.07
2	37.3	37.15	0.15
3	40.2	40.00	0.20
4	60.5	60.18	0.32
5	80.8	80.35	0.45

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = $\pm 0.7^\circ \text{C}$

Priyank Sharma

Calibrated By : Priyank Sharma
Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.