



The way to Technology
(16 Years of Excellence)

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)

Calibration Certificate

Page 1 of 1

11.06.2019
12.06.2019
11.06.2020
15.06.2019

Certificate Number : CAL/IN/S/503818

Calibrated For
M/s. Sanjeevan Medical Research Centre (P) Ltd.
24, Ansari Road, Daryaganj,
New Delhi - 02

Date of Request
Calibrated on
Valid Upto
Cert. Issue Date

Nomenclature		Description of Sample	
Range	Incubator (Thermostat)	Accuracy/Class	Not Mentioned
Resolution	0 to 110 ° C	Condition of UUC	Physically Ok
Make & Model No.	5° C	Calibration Performed at	At Site
Serial No.	Tanco	Calibration Procedure No.	CP/T - 03
Party ID Mark	SMRC / Lab / INCUB / 02	Reference IS Standard	IS:7358 - 1984
Location	Lab	Environmental Condition	Temperature = 25 ± 5° C Humidity = 40 to 70% RH

Master Equipment / Standard Used

Sr.N.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03 PMM/CTR/01	Tempsens Udaipur	CC284019000003882F	12.05.2020

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Thermal Calibration

CALIBRATION RESULTS

S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	30	29.50	0.50
2	35	34.13	0.87
3	40	38.85	1.15
4	50	48.50	1.50
5	80	78.03	1.97

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = $\pm 3^\circ \text{C}$

Priyank Sharma

Calibrated By : Priyank Sharma
Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.