

LAB MONTHLY SUMMARY

Lab Name DAY AND NIGHT LABORATORY Lab No 5284
 Month January Year 2020
 Constituent Chemistry II
 Group

Date of Result Entered : 15/01/2020

Date of Report Published : 08/02/2020

Sl. No.	Parameter	Method	Others (Any other instruments)	108	136.91	8.93	12.23	123.3 mg/dl	-0.85	2.35
1	GLUCOSE I	GOD-POD	OTHERS (Any other instruments)	29	37.40	4.82	1.80	39.4 mg/dl	0.00	0.67
2	UREA I	UREASE UV / GLDH	OTHERS (Any other instruments)	28	1.77	9.36	0.17	1.9 mg/dl	-0.32	0.06
3	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) / JENDRASSIK	OTHERS (Any other instruments)	29	4.70	6.34	0.30	4.8 g/dl	-0.70	0.11
4	T-PROTEIN I	BIURET - colorimetric	OTHERS (Any other instruments)	26	3.01	4.88	0.15	2.9 g/dl	-0.88	0.06
5	ALBUMIN I	BCG - colorimetric	OTHERS (Any other instruments)	27	5.37	11.71	0.63	5.9 mg/dl	-0.06	0.24
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	OTHERS (Any other instruments)	30	95.97	5.08	4.88	38.5 mg/dl	-0.88	1.78
7	CHOLESTEROL I	CHOD-PAP	OTHERS (Any other instruments)	26	124.96	7.40	9.24	104 mg/dl	-0.61	3.63
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	OTHERS (Any other instruments)							

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Page 1 of 1

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****