



ISO/IEC 17025:2005
Certificate No.: CC-2705

28/02/2020

CALIBRATION REPORT STATUS : PASSED

DESCRIPTION : Fixed Volume Pipette FV-500(500 μ l)

DEVICE ID : 16308202

CALIBRATION DATE : 28/02/2020 11:17 AM

Method ID : FV/500

TERMINAL ID : 19

ENVIRONMENTAL FACTORS

TEMP : 20.00 $^{\circ}$ C Z FACTOR : 1.0026 mm³/mg BARO. PRESSURE : 80.00 KPa REL. HUMIDITY : 60.00%

CALIBRATION STATISTICS

Vol (μ l)	No	Cum Wt (mg)	Vol (μ l)	Mean (μ l)	SD (μ l)	Inaccuracy E%		Imprecision CV%		Status
						Actual	Target	Actual	Target	
500.000	1	498.900	500.197	500.063	0.322	0.013	0.60	< 0.20	0.20	PASSED
	2	997.900	500.297							
	3	1496.300	499.696							



Volume	Above 10 μ l to 100 μ l	Above 100 μ l to 1000 μ l	Above 1 ml to 10 ml	Above 10 ml to 100 ml	Based on data in the records.
Uncertainty (k=2)	0.1 μ l	0.1 μ l	0.1 μ l	4 μ l	

- * Specifications conform to ISO:8655 standards.
- * Each instrument is individually calibrated on electronic balance.
- * 750 mmHg = 99.98 kPa.
- * Weight in mg or g.
- * Volume, Mean & S.D. in ml or μ l.

Reference standard
The instrument is calibrated using a standard electronic balance with calibration traceability to NPL.

The reported expanded uncertainty of measurement is calculated by multiplying the standard uncertainty of measurement by the coverage factor k=2, which for normal distribution corresponds to a coverage probability of approximately 95%.