

सिपेट : इंस्टीट्यूट ऑफ प्लास्टिक्स टेक्नोलॉजी (आई.पी.टी.)  
CIPET : INSTITUTE OF PLASTICS TECHNOLOGY (IPT)  
(रसायन एवं पेट्रोरसायन विभाग)  
(Department Of Chemicals & Petrochemicals)



CIPET सि पेट  
probe • perform • practice • Plastics

(रसायन एवं उर्वरक मंत्रालय, भारत सरकार)  
(Ministry Of Chemicals & Fertilizers, Govt. of India)

बी-२७, अमौसी इण्डस्ट्रियल एरिया, लखनऊ-२२६ ००८  
B - 27, Amausi Industrial Area, Lucknow - 226 008

ग्राहक सेवा/Customer Service: 8756992882 फोन/Phone : +91-522 2436227  
ई-मेल/e-mail:cipetlko2@gmail.com, cipetcalibration1314@gmail.com  
website: www.cipet.gov.in



CC2745

अंशांकन प्रमाण पत्र / Calibration Certificate



ULR No. : CC274520000001166F

CERTIFICATE NO. :	193/HG/0220	SHEET 1 OF 2
CALIBRATION OF :	Digital Thermo Hygrometer	
CALIBRATED ON :	27.02.2020	

1. Calibrated for : M/S Rani Laxmi Bai Hospital  
Lucknow

CALIBRATION DUE : 27.02.2021

2. Service Request No. & date:

3. Identification of Device Under Calibration (DUC)

- a) Range : 0 - 50 °C & 10 to 99 % RH  
b) Least Count : 0.1 °C & 1 % RH  
c) Calibrated Range : 15 to 40 °C & 30 to 85 % RH  
d) Make/Model : HTC-1  
e) Serial No. : RLB/TH-03  
f) Location : ---

4. Standard & Major equipment(s) used for calibration :

Standard(s)	Range/Leastcount	Limit of Accuracy/ Uncertainty in meas.	Traceability/ Validity
PRT Sensor with Digital Readout	-40 to 600 °C 0.001 °C	± 0.340 °C	016(1)/TC/0519 12.05.2020
Digital Hygrometer with Sensor (Rotronics)	0 to 100 % RH 0.1 % RH	± 0.87 % RH	HAS/MTL/25400910/T004 16.09.2020

5. Environmental Condition(s) : Temperature : 23±2°C

6. Calibration Procedure: CCC/CP-T/05 Relative Humidity: 50 ±10%

7. Calibration Results :

SHEET ENCLOSED

8. Inference :

- (a) The correction is positive except when indicated otherwise, it has to be algebraically added to the DUC value in order to get actual value.  
(b) The calibration certificate shall not be reproduced except in full without written approval of laboratory.  
(c) The above result related only to the instrument submitted in as received condition.  
(d) The DUC meets its stated requirements.  
(e) Label affixed on device

Checked by

Approved by  
Authorized Signatory

VIVEK KUMAR  
Sr. Technical Officer  
(Quality Manager)

सिपेट : इंस्टीट्यूट ऑफ प्लास्टिक्स टेक्नोलॉजी (आई.पी.टी.)  
CIPET : INSTITUTE OF PLASTICS TECHNOLOGY (IPT)  
(रसायन एवं पेट्रोरसायन विभाग)  
(Department Of Chemicals & Petrochemicals)



CIPET सि पे ट  
probe • perform • practice • Plastics

(रसायन एवं उर्वरक मंत्रालय, भारत सरकार)  
(Ministry Of Chemicals & Fertilizers, Govt. of India)

बी-२७, अमौसी इण्डस्ट्रियल एरिया, लखनऊ-२२६ ००८  
B - 27, Amausi Industrial Area, Lucknow - 226 008

ग्राहक सेवा/Customer Service: 8756992882 फोन/Phone : +91-522 2436227  
ई-मेल/e-mail:cipetlko2@gmail.com, cipetcalibration1314@gmail.com  
website: www.cipet.gov.in



CC2745

अंशांकन प्रमाण पत्र / Calibration Certificate



CERTIFICATE NO. :	193/HG/0220	SHEET 2 OF 2
CALIBRATION OF :	Digital Thermo Hygrometer	
CALIBRATED ON :	27.02.2020	CALIBRATION DUE: 27.02.2021

TEMPERATURE

Sl. No.	Scale	DUC Value	Std.Value (Avg)	Correction
a.	<sup>0</sup> C	15	15.2544	0.2544
b.	<sup>0</sup> C	20	20.2552	0.2552
c.	<sup>0</sup> C	25	25.0225	0.0225
d.	<sup>0</sup> C	30	30.2258	0.2258
e.	<sup>0</sup> C	35	35.2278	0.2278
f.	<sup>0</sup> C	40	40.2136	0.2136

\*\*Expanded Uncertainty of measurement =  $\pm 0.37$  <sup>0</sup>C  
with coverage factor  $k=2$  at 95.45 % of confidence level.

HUMIDITY

Sl. No.	Scale	DUC Value	Std.Value (Avg)	Correction
a.	%	30	29.40	-0.60
b.	%	50	49.60	-0.40
c.	%	70	69.40	-0.60
d.	%	80	79.60	-0.40
e.	%	85	84.10	-0.90

\*\*Expanded Uncertainty of measurement =  $\pm 1.82$  %  
with coverage factor  $k=2$  at 95.45 % of confidence level.

Checked by

Approved by  
Authorized Signatory

VIVEK KUMAR  
Sr. Technical Officer  
(Quality Manager)