



Calibration Certificate

This is to certify that the cobas c 111 is a fully automated analyzer with S.No: **15348**, At **Vijaya Diagnostics Pvt. Ltd, Nandyala** has been calibrated on 27-08-2019.

The calibration includes:

- Replacement of Filters
- Replacement of Lamp
- Checked Voltages
- Lubrication of Robotics
- Adjustments and Alignments of Probe
- Temperature Adjustments
- Performed Air/Water Calibration
- Reaction Disk: Range: 36.8°C to 37.2°C
Adjusted: 37°C
- Reagent Disk: Range: 8.0°C to 12.0°C
Adjusted: 10°C

All the temperatures and volumes drawn by pumps were checked and found they are OK.

Next calibration due on 26-08-2020.

For Roche Diagnostics India Pvt Ltd.,

B. Siva Kumar
Service Engineer
Hyderabad.