



TOSOH

TOSOH INDIA PVT.LTD.

Gebi Industrial Park, Building No. 'C', Sonale Village
 Bhiwandi, Thane -421 302, India, Tel.: +91-25 222 84100
 email: info@tosohindia.com, web: www.tosohindia.com

HLC-723 Report of Calibration Date		YY:	MM:	DD:
Variant Mode		No. 2019-08-02		
Laboratory Vijaya Diagnostic Address Nandyal Department Biochemistry	Signature	Section	Name Siedhant Das. Product Specialist	
Sign of Check ok: V replace: X adjustment: A repair R			cleaning :C	grease up: L
Machine Information		Measurement Conditions		
Instrument HLC-723 GX Instrument S/N (-0000) Sample Loader S/N (11418608 - NIA -) Needle type <input checked="" type="checkbox"/> standard <input type="checkbox"/> side-hole AUTO SAVE <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO Bar-code <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO	SYSTEM Sample FLOW Y1-SMP SYS-IN1 SYL-IN3	Ver. No. <input checked="" type="checkbox"/> whole blood <input type="checkbox"/> centrifuged <input type="checkbox"/> mixed Z1-SMP 3025 Y1-STAT — —	01.24.	
1. Alarm setting <input checked="" type="checkbox"/> paper-end, condition of printing <input checked="" type="checkbox"/> Column oven temperature	4. Pump <input checked="" type="checkbox"/> Check if the pump cam is not dirty <input checked="" type="checkbox"/> Leak from plunger-seal <input checked="" type="checkbox"/> Check if the diaphragm and plunger are not worn <input checked="" type="checkbox"/> Actions & leak of check-valves (Uptake & Purge) <input checked="" type="checkbox"/> Leak from drain-valve <input checked="" type="checkbox"/> Suction & line filters clogging <input checked="" type="checkbox"/> Pressure (MPa) w/o column 0.2 w/ column 6.5 MPa <input checked="" type="checkbox"/> Vacuum condition <input checked="" type="checkbox"/> Drain flush <input checked="" type="checkbox"/> Valves actions FLOW FACTOR 0.99 Measured flow 1.08 ml/min			
2. Sampling unit <input checked="" type="checkbox"/> Rotor-seal (Injection valve) <input type="checkbox"/> Stator-face (Injection valve) <input type="checkbox"/> Arm (Injection valve, check if loosened or not) <input type="checkbox"/> Sample-loop (Injection valve) <input type="checkbox"/> Rotor-seal (Rotary valve) <input type="checkbox"/> Round head screws at the coupling (Rotary valve) <input type="checkbox"/> Packing check leak and worn or not (Syringe-S) <input type="checkbox"/> Packing, check leak and worn or not (Syringe-L) <input type="checkbox"/> Dilution port & Needle wash block, clean or not <input type="checkbox"/> Leak from O-ring (P/N 017092) <input type="checkbox"/> Barcode reader, readable without error <input type="checkbox"/> End-marker, detection sensitivity <input type="checkbox"/> Positions of rack-holder, tube-holder & sample sensor <input type="checkbox"/> Actions of Z1 & Y1 axes and SY-S & SY-L <input type="checkbox"/> Actions of X1/X2, Y2 and Y3 axes, no loosened screw <input type="checkbox"/> Position of X1 pitch-sensor <input type="checkbox"/> Positions between the main body and loader <input type="checkbox"/> Needle positions (at sampling, at dilution, at STAT) <input type="checkbox"/> Sample suction positions Whole blood sample Aspiration check: 6 uL Diluted blood sample Aspiration check: 120 uL	5. Others <input checked="" type="checkbox"/> Quality of printed letters <input checked="" type="checkbox"/> Parameters settings <input checked="" type="checkbox"/> AC voltage L-N 224 N-E 0.3 L-E 225 <input checked="" type="checkbox"/> DC voltage 12V 11.8 24V 24 5V <input checked="" type="checkbox"/> Check of Smart Media <input checked="" type="checkbox"/> Clean up			
3. Detector Column oven Temperature 25 Allowed: +/- 2 <input checked="" type="checkbox"/> Check noise at buffer changes <input type="checkbox"/> Baseline stability <input checked="" type="checkbox"/> Detector adjustment Lamp Initial Intensity 205 Lamp Current Intensity 36.60W REF % 99.5	6. Receiving inspections <input checked="" type="checkbox"/> CALIB.std. 1: 5.85 % 2: 10.67 Calib. factors $Y = (1.1418)X + (0.3816)$ <input checked="" type="checkbox"/> CALIB 1, RT A1a: 0.24 A1b: 0.32 F: 0.42 L-A1c: 0.54 S-A1c: 0.67 A0: 1.02 <input checked="" type="checkbox"/> Cntl assigned values L: 5.6 ± 0.3% H: 9.9 ± 0.5% Observed Cntl values L: 5.3 % H: 9.9 <input checked="" type="checkbox"/> Total-Area <input checked="" type="checkbox"/> 700-3,000 <input type="checkbox"/> NG <input checked="" type="checkbox"/> Chromatogram <input checked="" type="checkbox"/> Good <input type="checkbox"/> NG			
	<input checked="" type="checkbox"/> Check if all the items on this sheet have been filled out.			
	7. Remarks Calibration due on 02-08-2020			