

CALIBRATION CERTIFICATE

NAME OF THE CUSTOMER	VIJAYA DIAGNOSTIC CENTER NANDYAL
-----------------------------	---

Make	Beckman Coulter
PRODUCT NAME	DxH520-Haematology Analyzer
SERIAL NO.	BC040592

This is to certify that the **DxH 520 Haematology analyzer** has been calibrated with calibrator (Lot.No-491911800, expiry-05/09/ 2019)

S.NO.	Procedures	parameter	REMARKS
01	Start up	WBC/RBC/HGB/PLT	PASS
02	Reproducibility	WBC/RBC/HGB/MCV/HCT/PLT	PASS
03	Carry over	WBC/RBC/HGB/PLT	PASS
04	QC	WBC/RBC/HGB/HCT/PLT	PASS
05	Calibration(Lot.no-491911800)	WBC/RBC/HGB/MCV/PLT	PASS

Date of calibration:05-08-2019	Due date for calibration:04/08/2020
---------------------------------------	--

For Beckman Coulter India private limited



Sunil Raj Vazrapu
Application Engineer