



# SYNDICATE DIAGNOSTICS PVT.LTD.

## CALIBRATION CERTIFICATE

This is to certify that START Max Sr.No. 12227644002121 installed at Vijaya Diagnostics Center, Nandyal. The Instrument has been serviced and calibrated on 05/08/2019. The instrument is working satisfactorily as per its specification.

Has been Serviced and calibrated, after calibration, following performance check has been carried out:

- 1) Parameter Test : Self check was performed to confirm the system Parameters
- 2) Temperature test : Temperature of 37 °C was confirmed in the Channels
- 3) Motorization test : The motorization test was performed to confirm the Accuracy of the motors  
Instrument is working satisfactorily as per the specifications

The Start Max was calibrated on : 05-08-2019

Next calibration is due on : 04-08-2020

For Syndicate Diagnostics Pvt. Ltd.

**CH Channakesava**

Service Engineer

Instrumentation department

REG. OFF: - 223, 1<sup>st</sup> FLOOR, BICHOON MARKET, KISHAN POLE BAZAR, JAIPUR 302003, RAJASTHAN

Sales office: - K-27, K- Block, Gali No. 6, Mahipalpur Extension New Delhi-110037.

Ph. No.011-26781258. E-mail: syndicatediagnostics@yahoo.co.in



# SYNDICATE DIAGNOSTICS PVT.LTD.


Equipment Name : Semi Automated Coagulation analyzer  
Model : StartMax  
Serial No. : 12227644002121

Make : DiagnosticaStago (France)

PM Check list for semi-automated coagulation analyzer

STEP	PROCEDURE	SPECIFICATION	OBSERVATION
1.	CLEANING		
a)	Cleaning the instrument surface	Instrument status	OK
b)	Cleaning each reagent storage position	Instrument status	OK
c)	Cleaning 4 measurement wells	Instrument status	OK
d)	Cleaning 16 incubation wells	Instrument status	OK
2.	Calibration		
a)	Temperature Calibration	37 °C ± 0.5 °C	37.0 °C
b)	Motorization calibration	Instrument status	OK
3.	Electrical safety check		
a)	Line-Neutral voltage	220-230 V	230 V
b)	Line-Ground voltage	220-230 V	230 V
c)	Ground-Neutral voltage	< 6 V	0.8 V

Next calibration is due on 04-08-2020

  
Service Engineer  
CH.Chennakesava

REG. OFF: - 223, 1<sup>st</sup> FLOOR, BICHOON MARKET, KISHAN POLE BAZAR, JAIPUR 302003, RAJASTHAN  
Sales office: - K-27, K- Block, Gali No. 6, Mahipalpur Extension New Delhi-110037.  
Ph. No.011-26781258. E-mail: syndicatediagnostics@yahoo.co.in