

Calibration Certificate

Following voltages have been checked & adjusted on DiaSys SYS 400 (Sl. No. SYS40190501) at
 LIFE PATH CARE, 323, 117/N269-B, TULSHI NAGAR, NEAR SHARMA DENTAL CLINIC,
 KAKADEO, KANPUR- 208025.

Calibration Report:

1. System voltages are in the specified range - Checked by Digital Multimeter are as below:

No.	Module	Observed Value	Acceptable Range
	Power supply board o/p	+24.1 VDC	+24.0 +/- 0.5 VDC
		+05.1 VDC	+05.0 +/- 0.2 VDC
2	Lamp voltage	+12.1 VDC	12.0 +/- 0.5 VDC

2. System Temperature values are in the specified range - Checked by Digital Thermometer are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Reaction Disk Temp	37.0 C	37.0 C +/- 0.3 C

3. Photometer Adjustment Values are within range as recommended:

S.No.	Module	Observed Value	Acceptable Range
1	ADC Count for all Wavelengths	10,100 to 10,700	8000 to 18,000

➤ All other mechanisms have been checked through field service program & Software, found OK.
 ➤ Performed calibration for End Point, Kinetic & Fixed Time Kinetic. Run control & sample. Results were satisfactory.
 ➤ Next Calibration will be performed on 14/07/2020.

For DiaSys Diagnostics India Pvt. Ltd.

Rizwan
 RIZWAN KILAN

Area Service Manager

Date: 15/07/2019

Registered Office
 Plot No A-821,
 TTC Industrial
 Area

DiaSysDiagnosticsIndiaPrivateLimited

CIN U74999MH2013PT