

LAB MONTHLY SUMMARY

Lab Name THE ERODE DIAGNOSTIC LABORATORY Lab No 7717  
 Month January Year 2020  
 Constituent Group Thyroid Hormones & Cortisol

Date of Result Entered : 19/01/2020

Date of Report Published : 11/02/2020

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	FREE THYROXIN	CLIA- chemiluminescence immunoassay / CMIA	SIEMENS Advia centaur,Immulite,Atellica Series	105	1.18	11.41	0.14	1.37 ng/dL	1.41	0.03
2	TSH	CLIA- chemiluminescence immunoassay / CMIA	SIEMENS Advia centaur,Immulite,Atellica Series	121	4.41	11.77	0.52	4.38 uIU/ml	-0.06	0.09

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*