

National Accreditation Board for Testing and Calibration Laboratories



(A Constituent Board Of Quality Council Of India)

SCOPE

Criteria: NABL 128 (Quality Assurance Scheme for Basic Composite Medical Laboratories -Entry Level)

Laboratory Name.: LABORATORY DEPARTMENT GENERAL HOSPITAL
RAJPIPLA

Certificate No.: QAS-BC-Entry Level-00001

Validity: 15-02-2020 to 14-02-2023

Page:

Last Amended on:



S. No.	Type of Sample examined/tested	Specific tests/examination performed	Standard(method), Principle, Methodology or Technique used
1	Serum	Renal Function Test -Creatinine	Modified Kinetic Jaffe
2	Serum	Renal Function Test - Urea	Urease
3	Serum	Liver Function Test - Total Bilirubin	Sulphanilic Acid
4	Serum	Liver Function Test - Direct Bilirubin	Diazotization
5	Serum	Liver Function Test - ALT (SGPT)	Optimized UV test
6	Serum	Liver Function Test - AST (SGOT)	Optimized UV test
7	Serum	HIV	Rapid Immunochromatography
8	Serum	HBsAg	Immuchromatography
9	Serum	WIDAL	Direct agglutination
10	Urine	WBC, RBC, Crystal, Cast	Photoelectric colour comparision - Strip / Microscopic
11	Whole Blood (EDTA)	CBC - Haemoglobin	Non Cyanmethemoglobin
12	Whole Blood (EDTA)	CBC - TLC (Total Leucocyte Count) / Total count of WBC	Electrical Impedence
13	Whole Blood (EDTA)	CBC - DLC (Differential Leucocyte Count) / Differential count of WBC	Electrical Impedence
14	Whole Blood (EDTA)	CBC - Platelet count	Electrical Impedance

This is annexure to 'Certificate of Accreditation' and does not require any signature.