

CERTIFICATE OF CALIBRATION

Customer Name : Department of Pathology AHM & Dufferin Hospital Kanpur Nagar.

Instrument Name : Medonic M series **Model :** M20 **Serial Number :** 30972

Calibration done on : 24/08/2019

Next Calibration due : 23/08/2020

This is to certify that abovementioned instrument has been calibrated for six CBC parameters i.e. RBC, MCV, PLT, MPV, HGB and WBC in open tube mode after verifying the background run for the machine. Calibration was verified by running controls (3 levels).

Calibration done by: Manish Kumar

Certificate issued by: Sunil Kumar Yadav
Sr. Application Specialist
POCT Services.

