

C-1, Shyam Gokul Appartment, Nr. Shyamal Cross Road,
Satellite, Ahmedabad - 380 015. Gujarat, India
Ph. : 079-26761712 Cell : (+91) 93761 48212, 90999 76212
Email : wecareenterprise@rediffmail.com, amitshah2612@gmail.com

WE CARE
ENTERPRISE



CALIBRATION CERTIFICATE

THIS IS TO CERTIFY THAT DXH 500 FIVE PART BLOOD CELL
COUNTER.SR.NO BBO20119 INSTALLED AT MODI CLINICAL LAB PVT
LTD -ANKLESHWAR HAS BEEN SUCCESSFULLY CALIBRATED ON
29/JUNE/2019 WITH DXH 500 CALIBRATOR LOT NO.: 4919111600
EXP. DATE: 05.07.2019

THE CALIBRATION DATA IS ENCLOSED WITH THIS CALIBRATION
CERTIFICATE.

FOR WE CARE ENTERPRISE


AMIT SHAH



Together No Cure

WE CARE ENTERPRISE

Lot Number: 491911600

Accept Date/Time: 29/06/2019 02:52 PM

Expiration Date: 05/07/2019

Source: BEC

| Run Date/Time | OPR. | Excl | Run | WBC | RBC | HGB | MCV | PLT | MPV |
|---------------------|---------|------|-----|------|------|-------|------|-------|------|
| 29/06/2019 02:41 PM | SERVICE | N | 1 | 8.90 | 4.19 | 14.00 | 92.3 | 262.5 | 8.56 |
| 29/06/2019 02:42 PM | SERVICE | N | 2 | 8.90 | 4.30 | 13.77 | 91.8 | 272.0 | 8.72 |
| 29/06/2019 02:44 PM | SERVICE | N | 3 | 9.01 | 4.27 | 13.94 | 92.0 | 262.4 | 8.73 |
| 29/06/2019 02:45 PM | SERVICE | N | 4 | 8.83 | 4.23 | 13.89 | 92.2 | 276.1 | 8.49 |
| 29/06/2019 02:46 PM | SERVICE | N | 5 | 8.92 | 4.21 | 13.94 | 92.1 | 272.4 | 8.45 |
| 29/06/2019 02:47 PM | SERVICE | N | 6 | 8.92 | 4.23 | 13.87 | 92.1 | 276.4 | 8.71 |
| 29/06/2019 02:48 PM | SERVICE | N | 7 | 9.18 | 4.26 | 13.87 | 92.0 | 271.5 | 8.66 |
| 29/06/2019 02:49 PM | SERVICE | N | 8 | 9.01 | 4.15 | 13.92 | 92.0 | 276.8 | 8.57 |
| 29/06/2019 02:50 PM | SERVICE | N | 9 | 9.10 | 4.23 | 13.93 | 92.0 | 258.5 | 8.67 |
| 29/06/2019 02:51 PM | SERVICE | N | 10 | 9.24 | 4.25 | 13.87 | 91.9 | 258.9 | 8.55 |

| | WBC | RBC | HGB | MCV | PLT | MPV |
|----------------|-------|-------|-------|-------|-------|-------|
| Mean | 9.00 | 4.23 | 13.90 | 92.0 | 268.8 | 8.61 |
| %CV | 1.44 | 0.95 | 0.43 | 0.11 | 2.75 | 1.16 |
| Target | 8.95 | 4.19 | 13.90 | 94.3 | 250.0 | 8.80 |
| Factor % Diff | -0.56 | -0.92 | 0.00 | 2.50 | -6.97 | 2.20 |
| Delta Diff | 0.05 | 0.04 | 0.00 | 2.3 | 18.8 | 0.19 |
| In-Use Factors | 1.065 | 0.870 | 1.133 | 1.082 | 0.990 | 1.000 |
| New Factor | 1.059 | 0.862 | 1.133 | 1.109 | 0.921 | 1.022 |
| Status | CAL | CAL | CAL | CAL | CAL | CAL |

| | | |
|-------------|-----------|---------------------|
| Diluent | 0028170 | 28/08/2019 |
| Lyse | 8300020 | 28/08/2019 |
| Cleaner | 8310066 | 27/09/2019 |
| Control | | |
| Calibration | 491810100 | 14/03/2018 09:28 AM |