



## Urine Routine Analysis

### ALL METHOD REPORT

Cycle - 9/2020

Date: 30/03/2020

Lab Code: 1838

Round No - 03

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	103	89	86%	1.010 - 1.020	1.010	With in consensus
pH	-	104	100	96%	6.0-7.0	7.0	With in consensus
Leukocytes	/µl	100	91	91%	1+ - 3+	3+	With in consensus
Nitrite	-	99	94	95%	Positive	.	-
Protein	mg/dl	104	90	87%	2+ - 4+	4+	With in consensus
Glucose	mg/dl	104	86	83%	3+ - 4+	4+	With in consensus
Ketones	-	103	86	83%	1+ - 3+	*4+	Outside consensus
Urobilinogen	mg/dl	104	90	87%	2+ - 4+	3+	With in consensus
Bilirubin	mg/dl	104	91	88%	3+	3+	With in consensus
Blood (Hemoglobin)	/µl	103	100	97%	3+ - 5+	5+	With in consensus

Legends	≥ 80% Consensus	(*) Outside Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	---------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Programme Director

Dr.Bandana Mehrotra

Checked By:

Page 1 of 4

