



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 1838

Cycle - 9/2020
Round No - 01

Date: 03/02/2020

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	84	78	93%	1.010 - 1.020	1.020	With in consensus
pH	-	85	77	91%	6.0	*7.0	Outside consensus
Leukocytes	/μl	83	82	99%	Negative	Negative	With in consensus
Nitrite	-	82	80	99%	Negative	Negative	With in consensus
Protein	mg/dl	85	83	98%	Negative	Negative	With in consensus
Glucose	mg/dl	84	83	99%	Negative	Negative	With in consensus
Ketones	-	85	84	99%	Negative	Negative	With in consensus
Urobilinogen	mg/dl	83	82	99%	Negative	Negative	With in consensus
Bilirubin	mg/dl	84	83	99%	Negative	Negative	With in consensus
Blood (Hemoglobin)	/μl	83	82	99%	Negative	Negative	With in consensus

Legends	≥ 80% Consensus	(*) Outside Consensus	(.) Not Reported	(#) Late Result Submission
---------	-----------------	-----------------------	------------------	----------------------------

Chief Coordinator

Sanjay Mehrotra
Dr.Sanjay Mehrotra

Programme Director

Bandana Mehrotra
Dr.Bandana Mehrotra

Checked By:

[Signature]

Page 1 of 4

