

LAB MONTHLY SUMMARY

Lab Name MODI CLINICAL LABORATORY PVT LTD Lab No 10934
 Month December Year 2019
 Constituent Chemistry II
 Group

Date of Result Entered : 18/12/2019

Date of Report Published : 04/01/2020

Sl.No	Analyte	Method	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
1	GLUCOSE I	HEXOKINASE	64	264.44	5.07	13.41	238 mg/dl	-1.97	3.35
2	UREA I	UREASE UV / GLDH	1475	133.68	16.16	21.61	143 mg/dl	0.43	1.13
3	CREATININE I	JAFFES KINETIC	2782	6.86	13.25	0.91	8.03 mg/dl	1.29	0.03
4	T.BILIRUBIN I	EVELYN-MALLOY	132	1.93	23.01	0.44	1.7 mg/dl	-0.52	0.08
5	T-PROTEIN I	BIURET	1745	5.11	12.50	0.64	4.3 g/dl	-1.27	0.03
6	ALBUMIN I	BCG	1732	3.09	10.66	0.33	2.7 g/dl	-1.19	0.02
7	URIC ACID I	ENZYMATIC	1777	8.70	14.44	1.26	9 mg/dl	0.24	0.06
8	CHOLESTEROL I	CHOD-PAP	3051	104.44	12.13	12.67	99 mg/dl	-0.43	0.46
9	TRIGLYCERIDE I	ENZYMATIC	1829	98.42	20.29	19.97	102 mg/dl	0.18	0.93

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

CHRISTIAN MEDICAL COLLEGE

CMC EXTERNAL QUALITY ASSURANCE SCHEME
2019 EQAS YEARLY SUMMARY

Group I : Laboratories who reported 10 to 12 months during 2019

Group II : Laboratories who reported 6 to 9 months during 2019

Group III : Laboratories who reported less than 6 months during 2019

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE I	Group III	351	0.99	A
2	UREA I	Group III	322	0.50	A
3	CREATININE I	Group III	351	0.84	A
4	T.BILIRUBIN I	Group III	278	0.40	A
5	T-PROTEIN I	Group III	211	1.28	B
6	ALBUMIN I	Group III	211	0.88	A
7	URIC ACID I	Group III	246	0.31	A
8	CHOLESTEROL I	Group III	342	0.69	A
9	TRIGLYCERIDE I	Group III	246	0.47	A

Only first 2 Groups are Graded

OMSDI	Grade	Remarks
0 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 3.00	C	Satisfactory
Above 3.00	D	Unacceptable

Your Lab No 10934

Constituent Group Chemistry II

Address

MODI CLINICAL LABORATORY PVT LTD ,
OPPOSITE STATE BANK OF INDIA, CHAUTA
NAKA ,
ANKLESHWAR ,
GUJARAT ,393001 .

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****