

HIN/MED/2020-2021/100658

22nd July 2019

CALIBRATION CERTIFICATE

This is to certify that the hematology analyser **YUMIZEN H500** bearing serial no: **904YOXH02339** installed at **Ayurshundra Hospital, Guwahati** was calibrated on 22nd July 2019.

Calibrator : ABX MINOCAL
Lot No. : CX436
Expiry Date : 05th Aug 2019

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 21st July 2020



Shrish Dixit

(Head – Products & Marketing)

For Horiba India Pvt. Ltd.