



CALIBRATION CERTIFICATE

Customer Details

Customer Name: Ayursundra Health Care Pvt. Ltd.

Address: Guwahati, Assam

Instrument Details

Instrument Type: Vitek 2 Compact 30

Serial No.: VK2C15036

S/w Version: 8.01

Calibration Details

Calibration Date: 29/12/2019

Next Calibration Due Date: 28/12/2020

- | | |
|--|------|
| 1. All the alignments are in the specified range | ✓Yes |
| 2. All the voltages are in the specified range | ✓Yes |
| 3. The Incubator & Cabinet temperatures are in the specified range | |
| a. The Incubator temperature is 35.5°C +/- 1°C | ✓Yes |
| 4. The Vacuum Pump check done | ✓Yes |
| a. The vacuum is 0.89PSIA+/- 0.04PSIA | ✓Yes |
| b. Vacuum leak test is OK | ✓Yes |
| 5. Optics Calibration done | ✓Yes |
| a. ADC values for both Tx1 & Tx3 is in range | ✓Yes |
| b. DAC values for both Tx1 & Tx3 is in range | ✓Yes |

Authorised Signatory

Engineer's Name: Dipankar Saha

Signature:

Date: 29/12/2019



bioMérieux India Pvt. Ltd.