

4/30/2020

External Quality Assurance Scheme - Print Monthly Summary

LAB MONTHLY SUMMARY

Lab Name EXCEL PRIMA DIAGNOSTICS
 Month March
 Constituent Group Chemistry I

Lab No 10940
 Year 2020

Date of Result Entered : 17/03/2020

Date of Report Published : 11/04/2020

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	253	298.09	3.80	11.34	301 mg/dl	0.26	1.43
2	UREA	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	257	55.11	4.31	2.37	53.6 mg/dl	-0.64	0.30
3	CREATININE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	258	2.84	4.44	0.13	2.7 mg/dl	-1.11	0.02
4	T.BILIRUBIN	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	259	1.90	9.39	0.18	1.9 mg/dl	0.00	0.02
5	T-PROTEIN	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	253	4.58	3.76	0.17	4.4 g/dl	-1.05	0.02
6	ALBUMIN	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	259	2.60	5.72	0.15	2.5 g/dl	-0.67	0.02
7	CALCIUM	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	250	7.61	4.08	0.31	7.3 mg/dl	-1.00	0.04
8	PHOSPHORUS	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	210	2.61	5.97	0.16	2.5 mg/dl	-0.71	0.02
9	URIC ACID	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	256	2.58	5.74	0.15	2.4 mg/dl	-1.22	0.02
10	CHOLESTEROL	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	243	91.05	5.37	4.89	90 mg/dl	-0.21	0.63
11	TRIGLYCERIDE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	243	117.75	6.75	7.95	118 mg/dl	0.03	1.02
12	HDL CHO	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	230	18.20	7.08	1.29	19 mg/dl	0.62	0.17
13	SODIUM	OTHERS (Any Other Principles / Methods)	OTHERS (Any other Analyzer)	258	130.27	3.70	4.82	131 mmol/L	0.15	0.60
14	POTASSIUM	OTHERS (Any Other Principles / Methods)	OTHERS (Any other Analyzer)	249	2.61	6.75	0.18	2.6 mmo/L	-0.06	0.02
15	CHLORIDE	OTHERS (Any Other Principles / Methods)	OTHERS (Any other Analyzer)	196	93.80	5.15	4.83	86.7 mmol/L	-1.47	0.69
16	AST	Dry Chemistry with PLP (P-5-P)	JOHNSON & JOHNSON DRYCHEM SERIES	243	194.49	6.20	12.06	198 U/L	0.29	1.55
17	ALT	Dry Chemistry with PLP (P-5-P)	JOHNSON & JOHNSON DRYCHEM SERIES	237	146.21	9.38	13.71	146 U/L	-0.02	1.78
18	ALP	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	256	125.56	9.38	11.77	130 U/L	0.38	1.47
19	AMYLASE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	126	66.72	12.08	8.06	77 U/L	1.27	1.44

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential
 Contact details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

4/30/2020

External Quality Assurance Scheme - Print Monthly Summary

LAB MONTHLY SUMMARY

Lab Name EXCEL PRIMA DIAGNOSTICS
Month March
Constituent Group HbA1c

Lab No 10940
Year 2020

Date of Result Entered : 17/03/2020

Date of Report Published : 11/04/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-ION EXCHANGE	OTHERS (Any other Analyzer)	51	6.53	11.02	0.72	6.1 %	-0.60	0.20

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential
Contact details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

***** End of Report *****

Name : Mr. PAVAN KUMAR (20PM30307) Ref. Doctor : DR. SELF
Age / Sex : 29 Year(s) / Male Client : Walk-in
Patient ID : P20000028533
Visit No. : 20PM30837

Registered Date : 18/01/2020 04:27 PM
Received On : 18/01/2020 04:28 PM
Reported On : 18/01/2020 07:25 PM




Report Type : Final

Test	Results	Units	Biological Reference Range
BIOCHEMISTRY			
Gamma-Glutamyl Transferase (GGT) - Serum	16.0	U/L	12-64
Gamma--Glutamyl-p-nitroanilide IFCC			

-- End of Report --

Kindly correlate clinically. If necessary discuss/repeat

Note: (*) Test Outsourced to NABL accredited Lab.


SAIKAUSALYA
BIOCHEMIST

TECHNICAL SUPERVISOR / TECHNICIAN
Completed By :

83/10, 8th Main Road, 3rd Block, Jayanagar, Bengaluru – 560 011 Page 1 / 1

☎ 080 46382333 | 4638.2222 ✉ admin@primadiagnostics.com

Home collection : ☎ +91 90609 19103 / +91 7618720723, Web : www.primadiagnostics.com

Name : Dr. PAVAN KUMAR K Ref. Doctor : DR. C/O.MAX LIFE INSURANCE Registered Date : 18/01/2020 08:53 AM
Age / Sex : 29.11 Year(s) / Male Client : MD INDIA MAX LIFE Received On : 18/01/2020 08:58 AM
Patient ID : P20000028412 Reported On : 18/01/2020 06:44 PM
Visit No. : 20PM30707



Report Type : Final

Test	Results	Units	Biological Reference Range
------	---------	-------	----------------------------

BIOCHEMISTRY


CAT-5 MAXLIFE MD INDIA

Gamma-Glutamyl Transferase (GGT) - Serum ?-glutamyl-p-nitroanilide Dry Chemistry	21.00	U/L	15-73
---	-------	-----	-------

-- End of Report --

Kindly correlate clinically. If necessary discuss/repeat

Note: (*) Test Outsourced to NABL accredited Lab.


Dr Krishna kumari
MBBS DCP
Consultant Pathologist
KMC NO:44988

TECHNICAL SUPERVISOR / TECHNICIAN **TOTAL X-RAY & LAB SOLUTIONS PVT LTD**

Completed By : nmswamy

Regd Office : # 72, Sukrutha Arcade, Sahukar Channaiah Road, Saraswathipuram, Mysuru, Karnataka, India - 570 009

☎ 0821 - 2545777 📞 +91 90084 00877 ✉ xlabsolutions2016@gmail.com

CIN - U85320KA2016PTC097807