



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3055

Distribution No.: 150-H

Month/Year: January/2020

Instrument ID: 11449

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 29-07-2020[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.7	4.45	9.15	9.6	0.0560	-0.66	0.25	0.1	0.0140	1.19
RBC x10 <sup>6</sup> /µl	1	3.52	3.46	6.98	7.02	0.0120	-0.21	0.06	0.02	0.0030	1.08
Hb g/dl	1	10.7	10.6	21.3	22.7	0.0490	-1.89	0.1	0.1	0.0120	0.00
HCT%	1	32.8	32.3	65.1	66.3	0.2350	-0.32	0.5	0.3	0.0370	0.54
MCV-fl	1	93.4	93.2	186.6	189.25	0.5880	-0.30	0.2	0.3	0.0470	-0.17
MCH-Pg	1	30.6	30.4	61	65	0.1510	-1.70	0.2	0.3	0.0330	-0.32
MCHC-g/dl	1	32.8	32.6	65.4	68.5	0.2730	-0.70	0.2	0.4	0.0350	-0.67
Plt. x10 <sup>3</sup> /µl	1	172	163	335	334	1.99	0.04	9	5	0.48	0.90
Retic %	2	1	0.8	1.8	3.5	0.12	-0.92	0.2	0.2	0.03	0.00

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=4 , Poly=47 L=3, E=3, Mono/Promono=3 , B1=4 P.M.=9, Mye=19, Meta=10, Other=basophils : 2%	Poly: 45-55, L: 2-7, nRBC/Eo/Mono/Blast/Pro : 0-5, Myelo: 15-25, Meta: 10-20, Baso: 0-3.		
RBC Morphology	3	Normocytic and Normochromic	Predominantly: Normocytic Normochromic, Anisocytosis. Moderate: Hypochromic, Mild: Microcytic.		
Diagnosis	3	Chronic myelogenous Leukemia-Chronic phase	Chronic Myeloid Leukemia ( Chronic Phase ) : CML-CP.		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	450	108	80.56	80.56	7.41	7.41	8.33	9.26
RBC x10 <sup>6</sup> /µl	1	450	108	84.26	84.26	3.7	5.56	12.04	9.26
Hb g/dl	1	450	108	87.04	90.74	9.26	0.93	3.7	8.33
HCT%	1	450	108	92.59	85.19	5.56	3.7	1.85	10.19
MCV-fl	1	450	108	89.81	92.59	7.41	1.85	2.78	5.56
MCH-Pg	1	450	108	89.81	84.26	7.41	8.33	2.78	7.41
MCHC-g/dl	1	450	108	93.52	83.33	5.56	10.19	0	5.56
Plt. x10 <sup>3</sup> /µl	1	450	108	85.19	91.67	4.63	1.85	10.19	6.48
ReticCount%	2	450	88	84.09	90.91	6.82	1.14	5.68	9.09
PS Assessment	3	450	97	Acceptable:90.2%,Warning Signal:7.8%,Unacceptable :2%					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----