



RML - Quality Assurance Program (RML - QAP)



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 1838

Cycle - 9/2020

Date: 30/07/2020

Round No - 07

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	92	77	84%	1.010 - 1.020	1.020	With in consensus
pH	-	93	74	80%	6.0	6.0	With in consensus
Leukocytes	/ μ l	91	83	91%	Negative	Negative	With in consensus
Nitrite	-	85	82	96%	Negative	Negative	With in consensus
Protein	mg/dl	93	87	94%	Negative	Negative	With in consensus
Glucose	mg/dl	93	89	96%	Negative	Negative	With in consensus
Ketones	-	93	89	96%	Negative	Negative	With in consensus
Urobilinogen	mg/dl	93	89	96%	Negative	Negative	With in consensus
Bilirubin	mg/dl	93	89	96%	Negative	Negative	With in consensus
Blood (Hemoglobin)	/ μ l	92	88	96%	Negative	Negative	With in consensus

Legends	\geq 80% Consensus	(*) Outside Consensus	(.) Not Reported	(#)Late Result Submission
---------	----------------------	-----------------------	------------------	---------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Programme Director

Dr.Bandana Mehrotra

Checked by

Page 1 of 4



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines) 4077180, Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com

Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence



RML - Quality Assurance Program (RML - QAP)



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 1838

Cycle - 9/2020
Round No - 06

Date: 30/07/2020

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	94	82	87%	1.010 - 1.020	1.010	With in consensus
pH	-	95	91	96%	6.0-7.0	7.0	With in consensus
Leukocytes	/μl	91	82	90%	1+ - 3+	3+	With in consensus
Nitrite	-	88	80	91%	Positive	Positive	With in consensus
Protein	mg/dl	95	87	92%	2+ - 4+	4+	With in consensus
Glucose	mg/dl	94	88	94%	2+ - 4+	4+	With in consensus
Ketones	-	95	89	94%	1+ - 4+	4+	With in consensus
Urobilinogen	mg/dl	95	81	85%	2+ - 4+	3+	With in consensus
Bilirubin	mg/dl	94	82	87%	3+	3+	With in consensus
Blood (Hemoglobin)	/μl	94	92	98%	3+ - 5+	5+	With in consensus

Legends	≥ 80% Consensus	(*) Outside Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	---------------------------

Note:- Your Specific Gravity value is with in range.

Chief Coordinator

Dr.Sanjay Mehrotra

Programme Director

Dr.Bandana Mehrotra

Checked By:

Page 1 of 4



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines) 4077180, Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com

Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence