

Dear NABL

In between because of CAA and Covid19, sample dispatch has been very erratic/not been delivered to us because of all the lockdowns. Logistics problems have been too much because of the location of our state. We have received only the February 2020 sample and have attached that report. Also attached is the screenshot of the email showing application to the EQAP program. The EQAP provider has mentioned in their website that the May 2020 samples will not be sent due to the pandemic



Dr. Juri B. Kalita

Group Head—Quality

Ayursundra Superspecialty Hospital

Date: 12/08/2020



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 1696

Distribution No.: 150-D

Month/Year: February/2020

Instrument ID: HORIBA YUMIZEN H500

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 29-07-2020[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /μl	1	0.27	0.25	0.52	3.8	0.0220	-5.06	0.02	0.05	0.0030	-0.40
RBC x10 ⁶ /μl	1	3.76	3.66	7.42	7.77	0.0030	-2.01	0.1	0.03	0.0020	1.89
Hb g/dl	1	11.9	11.9	23.8	24.4	0.0180	-1.35	0	0.1	0.0070	-1.35
HCT%	1	32.6	32	64.6	71.4	0.1530	-1.46	0.6	0.3	0.0180	1.01
MCV-fl	1	87.2	86.8	174	184.2	0.3120	-1.00	0.4	0.2	0.0180	0.67
MCH-Pg	1	32.6	31.6	64.2	62.8	0.0550	0.99	1	0.3	0.0160	3.15
MCHC-g/dl	1	37.3	36.5	73.8	68.1	0.1340	1.29	0.8	0.3	0.0160	2.25
Plt. x10 ³ /μl	1	148	142	290	245	2.52	1.31	6	4	0.26	0.54
Retic %	2	5.1	5	10.1	6.2	0.25	1.02	0.1	0.2	0.01	-0.34

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs= , Poly=56 L=23, E=00, Mono/Promono=01 , B1=14 P.M.=00, Mye=03, Meta=03, Other=	Poly: 55-65, L: 2-10, nRBC/Eo/Mono/Pro: 0-5, Myelo: 10-20, Meta: 10-20, Baso: 0-2		
RBC Morphology	3	microcytic, hypochromic, anisocytosis, poikilocytosis	Predominantly: Normocytic Normochromic. Moderate: Anisocytosis. Mild: Microcytic.		
Diagnosis	3	AML	Chronic Myeloid Leukemia (Chronic Phase) : CML-CP		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	450	323	88.85	60.06	4.33	38.08	7.12	1.86
RBC x10 ⁶ /µl	1	450	323	87.31	87	6.19	5.88	6.81	7.43
Hb g/dl	1	450	323	84.83	94.12	8.36	2.79	6.81	3.1
HCT%	1	450	323	96.9	91.02	1.55	3.72	1.55	5.26
MCV-fl	1	450	323	99.38	88.54	0.31	6.81	0.31	4.95
MCH-Pg	1	450	323	87.31	93.5	7.74	2.79	5.26	3.72
MCHC-g/dl	1	450	323	98.14	89.78	1.24	6.19	0.62	3.72
Plt. x10 ³ /µl	1	450	323	91.64	89.78	5.57	5.57	3.1	4.95
ReticCount%	2	450	265	94.72	87.17	4.53	11.32	0.75	3.4
PS Assessment	3	450	310	Acceptable:95.5%, Warning Signal:4.5%, Unacceptable :0%					

Comments:

- 1). Among Lab (EQA) : Wrongly Reported PS Diagnosis, remaining results acceptable
- 2). Within Lab (IQA) : Difference in the CBC measurement values for MCH unacceptable, may be due to random/human error.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)
 Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Report authorized by,



Dr. Seema Tyagi (Prof.)
 PT Co-ordinator: ISHTM-AIIMS-EQAP
 Department of Hematology, AIIMS, New Delhi

-----End Of Report-----