

LAB MONTHLY SUMMARY

Lab Name **ABC MICRO LAB** Lab No **8232**
 Month **July** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : 01/07/2020

Date of Report Published : 14/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	574	390.61	11.17	43.64	426.5 mg/dl	0.82	3.64
2	UREA I	UREASE UV / GLDH	Transasia / Erba	518	91.03	13.34	12.14	101.6 mg/dl	0.87	1.07
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Transasia / Erba	461	1.00	15.82	0.16	1 mg/dl	0.00	0.01
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	458	4.04	12.45	0.50	4.1 mg/dl	0.12	0.05
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	422	5.55	9.07	0.50	5.5 g/dl	-0.10	0.05
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	423	3.29	9.34	0.31	3.2 g/dl	-0.29	0.03
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	471	7.59	12.33	0.94	8.5 mg/dl	0.97	0.09
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	501	104.73	12.51	13.10	99 mg/dl	-0.44	1.17
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	501	203.31	14.18	28.83	223.7 mg/dl	0.71	2.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****