



CALIBRATION CERTIFICATE CENTRIFUGE

Instrument Details		
Instrument Name	Centrifuge	
Model/ Serial No	Neya 16	
Calibration Date	18-08-2020	
Calibration Due Date	19-08-2021	
Reference Standard Details		
Reference Instrument Name	Tachometer	
Model No	Lutron/2234C	
Serial No	L180349	
Calibration Date	17-08-2020	
Calibration Due Date	18-08-2021	
Traceability Certificate No	CC223920600022818 F (Validity- 16-08-2021)	
Calibration Done For	AHM & Dufferin Female District Hospital, Kanpur	
Method of calibration	Calibration done using calibrated digital tachometer	
Results		
Instrument Reading (RPM)	Tachometer Reading (RPM)	% Variation
1500	1502	1.96
2000	2002	2.63
2500	2503	2.38
3000	3006	3.01
3500	3507	1.80
4000	4008	1.23

Acceptance Criteria: Less than $\pm 5\%$

Reference: SMILE Johns Hopkins University Baltimore, MD USA
Centrifuge Calibration Verification - Guidelines


Calibration Done By:
Mr. Shashank Yadav
POCT Services


Name & Signature:
Lab Incharge
डा० सुरभि सक्सेना
M.B.B.S., D.C.P.
पैथालाजिस्ट
ए०एच०एम० इफारिन अस्पताल
कानपुर नगर