

CERTIFICATE OF CALIBRATION

Ref. No.: MMH014
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: LAB STORAE FRIDGE
Make/Model: WHIRLPOOL/23 CLASSIC MASTER MIND
Sl. No: ...


Condition at receipt: Functional Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020

Date of Issue :

Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH014
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	LAB STORAE FRIDGE	S No.	...
Make/Model No.	WHIRLPOOL/23 CLASSIC MASTER MIND	Location	LAB
ID No.	MMH/LAB/FRI-1/001	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	4	4.2	±2	P
2	4	4.3	±2	P
3	4	4.2	±2	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH015
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: LAB STORAE FRIDGE
Make/Model: WHIRLPOOL/ICE MAGIC
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020

Date of Issue :

Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH015
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	LAB STORAE FRIDGE	S No.	...
Make/Model No.	WHIRLPOOL/ICE MAGIC	Location	LAB
ID No.	MMH/LAB/FRI-2/002	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION


Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	4	4	±2	P
2	4	4	±2	P
3	4	4.1	±2	P


REMARKS:-

- ❖ This result of calibration refers to only to the particular-items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By



Approved By
(QM)



(SE)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH016
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: LAB STORAE FRIDGE
Make/Model: PENPOLE/COMPOSAFE
Sl. No: ...

Condition at receipt: **Functional** **Calibrated at** : **Onsite**
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020

Date of Issue :

Approved Signatory



Cyrix Healthcare Pvt. Ltd.
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH016
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	LAB STORAE FRIDGE	S No.	...
Make/Model No.	PENPOLE/COMPOSAFE	Location	LAB
ID No.	MMH/LAB/FRI-3/003	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO. HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	4	4.1	±2	P
2	4	3.9	±2	P
3	4	3.9	±2	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/01	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME:	MICRO PIPETTE	SI.NO:
MAKE/MODEL.NO:	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-1/001

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-2/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC



ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	1000	999.6	999.6	999.7	999.4	999.5	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/02	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:
MAKE/MODEL.NO:	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-2/002

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-2/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%



CALIBRATION RESULTS							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	500	500.3	500.4	500.4	500.3	500.3	P

EMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By


(SE)



Approved By

(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/03	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	...
MAKE/MODEL.NO:	...	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-3/003

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-2/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ)	MEASURED READING (μ)					STATUS
		I	II	III	IV	V	
1	200	200.8	200.8	200.7	200.9	200.8	P

EMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By



Approved By

Cyrix Healthcare Pvt. Ltd.
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/04	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	...
MAKE/MODEL.NO:	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-4/004

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-2/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%



CALIBRATION RESULTS							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	99.6	99.7	99.7	99.5	99.7	P

EMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By


(SE)

Approved By

(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph: 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in

CYRIX™

HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/05	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	...
MAKE/MODEL.NO:	...	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-5/005

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-2/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC


ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%


CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	20	20.1	19.9	19.7	19.8	19.9	P

EMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By





Approved By

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/06	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	...
MAKE/MODEL.NO:	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-6/006

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-2/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	10	10.2	10.1	10.3	10.4	10.1	P

EMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE)

Approved By

(QM)



Cyrix Healthcare Pvt. Ltd.
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/07	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:
MAKE/MODEL.NO:	...	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-7/007

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2020	21/03/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	5	5.01	4.88	5.0	5.1	5.01	P
2	10	10.0	9.9	9.8	10.1	9.9	P
3	50	50	50.1	50.2	50.1	50.1	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

Approved By



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MORRIS MATHIAS HOSPITAL		
ADDRESS	355, K.P. Road, Nagercoil - 629001.	DATE OF CALIBRATION	20-08-2020
CERTIFICATE NO:	CHPL/PC/MMH/08	VALID UP TO	19-08-2021
		DATE OF ISSUE	21-08-2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	...
MAKE/MODEL.NO:	...	VISUAL INSPECTION	OK
LOCATION :	ASSET ID NO :	MMH/LAB/MP-8/008

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2020	21/03/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6 ^o C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	10	10.0	9.9	9.8	10.1	9.9	P
2	50	50	50.1	50.2	50.1	50.1	P
3	100	100	100.3	100.3	100.3	100.2	P

REMARKS:-
 result of calibration refers to only to the particular items submitted for calibration.
 above result is valid at the time of and under the stated conditions measurement.
 ibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific
 use only and should not used for commercial.

Calibrated By



Approved By

Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph: 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in

CYRIXTM
 HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH013
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: WEIGEING MACHINE
Make/Model:
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MM013
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	WEIGHING MACHINE	S No.	...
Make/Model No.	MMH/LAB/PB-12/030	Location	LAB
ID No.	...	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used	
Equipment Name	MASS
Make	...
Model/SR No.	...
Calibration Validity	22/01/2021
Calibration By	FCRI

Calibration Results				
SL NO	SET READING (Kg)	MEASURED READING (Kg)	TOLERANCE (Kg)	STATUS
1	10	10.3	±1	P
2	20	20.4	±1	P
3	60	60.4	±1	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE)

Approved By

(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTHCARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH001
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: DRY BATH INCUBATOR
Make/Model: SPAIN
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH001
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	DRY BATH INCUBATOR	S No.	...
Make/Model No.	SPAIN	Location	LAB
ID No.	MMH/LAB/DBI-1/001	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	35	35.4	±2	P
2	45	45	±2	P
3	60	60	±2	P


REMARKS:-

- ❖ This result of calibration refers to only to the particular-items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION


Ref. No.: MMH002
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: LAB INCUBATOR
Make/Model: ADCO-BACT
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020
Date of Issue :




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH002
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	LAB INCUBATOR	S No.	...
Make/Model No.	ADCO-BACT	Location	LAB
ID No.	MMH/LAB/INCU-1/019	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	35	35.1	±2	P
2	45	45	±2	P
3	60	60.3	±2	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration. —
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By

(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION


Ref. No.: MMH003
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: VERTICAL AUTOCLAVE
Make/Model: GOLEY
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020
Date of Issue :




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH003
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	VERTICAL AUTOCLAVE	S No.	...
Make/Model No.	GOLEY	Location	LAB
ID No.	MMH/LAB/AUTO-1/028	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	65	65	±2	P
2	85	86	±2	P
3	120	120	±2	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTHCARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH004
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: VERTICAL AUTOCLAVE
Make/Model: GOLEY
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH004
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	VERTICAL AUTOCLAVE	S No.	...
Make/Model No.	GOLEY	Location	LAB
ID No.	MMH/LAB/AUTO-2/021	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER
Make	HTC-2	HTC
Model/SR No.	...	IRX-64/20170804051
Calibration Validity	23/09/2020	13/01/2021
Calibration By	STIC	PERFECT CALLIBRATION

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	65	65	±2	P
2	85	85	±2	P
3	120	120	±2	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: MMH005
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: CENTRIFUGE
Make/Model: ELECTRA
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020
Date of Issue :


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH005
		Date of Calibration	20-08-2020
		Valid Up to	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	CENTRIFUGE	S No.	...
Make/Model No.	ELECTRA	Location	LAB
ID No.	MMH/LAB/CEN-3/005	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/07089699	...
Calibration Validity	16/01/2021	23/09/2020
Calibration By	STIC	STIC

Environmental Condition			
Temperature	27°C	Relative Humidity	59%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE(°C)	STATUS
1	1000	957	±5%	P
2	2000	1965	±5%	P
3	2500	2459	±5%	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)



Approved By

(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH006
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: CENTRIFUGE
Make/Model: ELECTRA LAB
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH006
		Date of Calibration	20-08-2020
		Valid Up to	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	CENTRIFUGE	S No.	...
Make/Model No.	ELECTRA LAB	Location	LAB
ID No.	MMH/LAB/CEN-2/022	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/07089699	...
Calibration Validity	16/01/2021	23/09/2020
Calibration By	STIC	STIC

Environmental Condition			
Temperature	27°C	Relative Humidity	59%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE(°C)	STATUS
1	1000	958	±5%	P
2	2000	1987	±5%	P
3	2500	2573	±5%	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring _____ scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)



Approved By

(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION


Ref. No.: MMH007
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: CENTRIFUGE
Make/Model: REMI
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRUXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH007
		Date of Calibration	20-08-2020
		Valid Up to	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	CENTRIFUGE	S No.	...
Make/Model No.	REMI	Location	LAB
ID No.	MMH/LAB/BCF-1/014	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/07089699	...
Calibration Validity	16/01/2021	23/09/2020
Calibration By	STIC	STIC

Environmental Condition			
Temperature	27°C	Relative Humidity	59%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE(°C)	STATUS
1	1000	987	±5%	P
2	2000	1989	±5%	P
3	2500	2498	±5%	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)



Approved By

(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com


CERTIFICATE OF CALIBRATION


Ref. No.: MMH008
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: MICROSCOPE
Make/Model: ABOMED
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH008
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	MICROSCOPE	S No.	...
Make/Model No.	ABOMED	Location	LAB
ID No.	MMH/LAB/MIC-4/007	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYSER	THERMO HYGROMETER	
Make	FLUKE	HTC-2	
Model/SR No.	ESA612/2959012	...	
Calibration Validity	17/01/2021	23/09/2020	
Calibration By	STQC	STIC	

Environmental Condition			
Temperature °C	27°C	Relative Humidity %	54%

Calibration Results	
ELECTRICAL SAFETY REPORT ATTACHED	

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com


CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES


TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.3	3	0	V	P
4	EQUIPMENT CURRENT	0.4	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.4	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	167	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	316	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.3	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.2	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

*NOTE (NC-NORMAL CONDITION, OC-OPEN CONDTION)

Service Engineer




Quality Manager



Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph: 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in

CYRIXTM
 HEALTHCARE PVT LTD

www.cyrixhealthcare.com



CERTIFICATE OF CALIBRATION


Ref. No.: MMH009
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: MICROSCOPE
Make/Model: C POULO
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH009
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	MICROSCOPE	S No.	...
Make/Model No.	C POULO	Location	LAB
ID No.	MMH/LAB/MIC-2/020	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYSER	THERMO HYGROMETER	
Make	FLUKE	HTC-2	
Model/SR No.	ESA612/2959012	...	
Calibration Validity	17/01/2021	23/09/2020	
Calibration By	STQC	STIC	

Environmental Condition			
Temperature °C	27°C	Relative Humidity %	54%

Calibration Results	
ELECTRICAL SAFETY REPORT ATTACHED	


REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.3	3	0	V	P
4	EQUIPMENT CURRENT	0.4	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.4	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	167	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	316	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.3	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.2	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	NA	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

* NOTE (NC-NORMAL CONDITION, OC-OPEN CONDTION)

Service Engineer

Quality Manager

Cyrix Healthcare Pvt. Ltd.,

30/641 B Petta Junction, Poonithura, Kochi-38.

Ph: 97455 55073

Email: sales.kl@cyrix.in

service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: MMH010
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: BINOCULAR MICROSCOPE
Make/Model: LAB TECH
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH010
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	BINOCULAR MICROSCOPE	S No.	...
Make/Model No.	LAB TECH	Location	LAB
ID No.	MMH/LAB/MIC-5/023	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYSER	THERMO HYGROMETER	
Make	FLUKE	HTC-2	
Model/SR No.	ESA612/2959012	...	
Calibration Validity	17/01/2021	23/09/2020	
Calibration By	STQC	STIC	

Environmental Condition			
Temperature °C	27°C	Relative Humidity %	54%

Calibration Results	
ELECTRICAL SAFETY REPORT ATTACHED	

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)

Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com


CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES


TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	227	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	226	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.4	3	0	V	P
4	EQUIPMENT CURRENT	0.4	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.8	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	184	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	293	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.2	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.1	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

* NOTE (NC -NORMAL CONDITION, OC-OPEN CONDTION)

Service Engineer




Quality Manager



Cyrix Healthcare Pvt. Ltd.,

30/641 B Petta Junction, Poonithura, Kochi-38.

Ph: 97455 55073

Email: sales.kl@cyrix.in

service.kl@cyrix.in

CYRIX™

HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION


Ref. No.: MMH011
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: HAEMOSTASIS ANALYSER
Make/Model: XF-2.0
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 21-08-2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH011
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	HAEMOSTASIS ANALYSER	S No.	...
Make/Model No.	XF-2.0	Location	LAB
ID No.	MMH/LAB/HA-2/024	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYSER	THERMO HYGROMETER	
Make	FLUKE	HTC-2	
Model/SR No.	ESA612/2959012	...	
Calibration Validity	17/01/2021	23/09/2020	
Calibration By	STQC	STIC	

Environmental Condition			
Temperature °C	27°C	Relative Humidity %	54%

Calibration Results	
ELECTRICAL SAFETY REPORT ATTACHED	

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)

Approved By

(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com


CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES


TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.4	3	0	V	P
4	EQUIPMENT CURRENT	0.6	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.4	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	182	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	289	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.1	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.4	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

* NOTE (NC -NORMAL CONDITION, OC-OPEN CONDITON)

Service Engineer



Quality Manager



Cyrix Healthcare Pvt. Ltd.,

30/641 B Petta Junction, Poonithura, Kochi -38.

Ph: 97455 55073

Email: sales.kl@cyrix.in

service.kl@cyrix.in

CYRIXTM

HEALTH CARE PVT LTD

www.cyrixhealthcare.com



CERTIFICATE OF CALIBRATION

Ref. No.: MMH012
Applicant: MORRIS MATHIAS HOSPITAL
Device Calibrated: ROTATOR
Make/Model: LABLINE
Sl. No: ...

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 20-08-2020
Calibration due date : 19-08-2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

21-08-2020
Date of Issue :


Approved Signatory



Cyrix Healthcare Pvt. Ltd.
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MORRIS MATHIAS HOSPITAL 355, K.P. Road, Nagercoil - 629001.	Certificate No.	MMH012
		Date of Calibration	20-08-2020
		Valid Upto	19-08-2021
		Date of Issue	21-08-2020

Instruments Details			
Instrument Name	ROTATOR	S No.	...
Make/Model No.	LABLINE	Location	LAB
ID No.	MMH/LAB/RO-1/013	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYSER	THERMO HYGROMETER	
Make	FLUKE	HTC-2	
Model/SR No.	ESA612/2959012	...	
Calibration Validity	17/01/2021	23/09/2020	
Calibration By	STQC	STIC	


Environmental Condition			
Temperature °C	27°C	Relative Humidity %	54%

Calibration Results	
ELECTRICAL SAFETY REPORT ATTACHED	


REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.6	3	0	V	P
4	EQUIPMENT CURRENT	0.2	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.7	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	192	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	291	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.2	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.7	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

* NOTE (NC -NORMAL CONDITION, OC-OPEN CONDITON)

Service Engineer



Quality Manager



Cyrix Healthcare Pvt. Ltd.,

30/641 B Petta Junction, Poonithura, Kochi-38.

Ph: 97455 55073

Email: sales.kl@cyrix.in
service.kl@cyrix.in



CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com