



ISO/IEC 17025:2005
Certificate No.: CC-2705

3/03/2020

CALIBRATION REPORT STATUS : PASSED

DESCRIPTION : Variable Volume Pipette VV-1000(100-1000 µl)
DEVICE ID : 16411310
CALIBRATION DATE : 3/03/2020 5:04 PM
Method ID : VV/100-1000
TERMINAL ID : 20

ENVIRONMENTAL FACTORS

TEMP : 20.00 °C **Z FACTOR :** 1.0026 mm³/mg **BARO. PRESSURE :** 80.00 KPa **REL. HUMIDITY :** 60.00%

CALIBRATION STATISTICS

Vol (µl)	No	Cum Wt (mg)	Vol (µl)	Mean (µl)	SD (µl)	Inaccuracy E%		Imprecision CV%		Status	
						Actual	Target	Actual	Target		
100.000	1	100.000	100.260	100.427	0.209	0.427	6.00	< 2.00	2.00	PASSED	
	2	200.100	100.360								
	3	300.500	100.661								
500.000	1	498.300	499.596	499.429	0.474	0.114	1.20	< 0.40	0.40	PASSED	
	2	996.800	499.796								
	3	1494.400	498.894								
1000.000	1	1000.100	1002.700	1002.433	0.306	0.243	0.60	< 0.20	0.20	PASSED	
	2	2000.000	1002.500								
	3	2999.500	1002.099								

Volume	Above 10 µl to 100 µl	Above 100 µl to 1000 µl	Above 1 ml to 10 ml	Above 10 ml to 100 ml	Based on data in the records
Uncertainty (k=2)	0.1 µl	0.1 µl	0.1 µl	4 µl	

Reference standard
 The instrument is calibrated using a standard electronic balance with calibration traceability to NPL.

The reported expanded uncertainty of measurement is calculated by multiplying the standard uncertainty of measurement by the coverage factor k=2, which for normal distribution corresponds to a coverage probability of approximately 95%.

* Specifications conform to ISO:8655 standards.
 * Each instrument is individually calibrated on electronic balance.
 * 750 mmHg = 99.98 KPa
 * Weight in mg or g.
 * Volume, Mean & S.D. in ml or µl.

