



Date:  
Effective Date:

13-08-2020  
13-08-2020...  
UNMATCHED SERVICE  
NO 179...

## Certificate of Calibration

**Customer Name: OncQuest Laboratories Ltd**

**Model : Automated Hematology Analyzer Sysmex XP-100**

**Serial No. : B4039**

**Calibration Done Date: 13.8.20**

**Next Calibration Due Date On or Before: 12.08.2021**

**Lab In-charge: . Mr.Deepak Kumar**

*This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.*

*The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.*

Calibration at site performed by  
Engineer Name:Suman Kumar Saha  
Designation:Sr.Service Engineer  
Transasia Bio-Medicals Ltd  
Location:Kolkata

Encl:

1. Certificate of Inspection
2. Assay Sheet of Calibrator SCS-1000
3. Printouts
4. Traceability & Uncertainty document





UNMATCHED SERVICE

Date: 13-08-2020  
Effective Date: 13-08-2020

## Certificate of Inspection

1. Model: Automated Hematology Analyzer Sysmex XP – 100
2. Serial No.: B4039
3. Calibration Date: 13-08-2020
4. Material used: SCS-1000 (Lot No. 0196 0525, Expiry date: -16-Aug-2020)

By comparing your data to the results of the standard counters in Sysmex Corporation, the calibration for CBC 5 parameters using the measurement standard material (SCS-1000) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.

Technical Service Department  
Transasia Bio-Medicals Ltd

