

CALIBRATION CERTIFICATE


Identification of Instrument:	Refrigerator Thermometer		
Make/Model No:	Western/AG-1298	Serial no:	31749161200270
Calibration Done on:	14/09/2020	Calibration Due Date:	13/09/2021
Reference Standard:	Calibrated Thermometer		
Make/Model No:	Mextech/Multi Thermometer	Serial no:	M20883197
Calibration Done on:	28/08/2020	Calibration Due Date:	28/08/2021
Calibrated For:	District Women Hospital, Pathology Lab, Pratapgarh		
Certificate no:	81(01)/TG/0320		


Temp. of Ref Thermometer	2°C	4°C	6°C	8°C
S. No.	Temp.reading of Device	Temp.reading of Device	Temp.reading of Device	Temp.reading of Device
1	2.01	4.16	6.1	8.21
2	2.07	4.19	6.23	8.34
3	2.05	4.17	6.19	8.41
4	2.09	4.11	6.18	8.13
Avg	2.055	4.1575	6.175	8.2725

Calibration Result:

SN.	Scale	Standard Value	Device Value (Avg)	Correction
1	Degree Celcius	2	2.04	0.04
2	Degree Celcius	4	4.12	0.12
3	Degree Celcius	6	6.2	0.2
4	Degree Celcius	8	8.2	0.2

Reference: SMILE Johns Hopkins University Baltimore, MD USA


Calibration Done By
Mr. Shashank Yadav
Quality Specialist
POCT Services


Name & Signature
Lab Incharge महिला चिकित्सालय
प्रतापगढ़