



# CALIBRATION CERTIFICATE CENTRIFUGE

Instrument Details		
Instrument Name	Centrifuge	
Make / Model	Remi / R-8C	
Serial Number / UID	32115	
Calibration Date	14-09-2020	
Calibration Due Date	13-09-2021	
Reference Standard Details		
Reference Instrument Name	Tachometer	
Model No	Lutron/2234C	
Serial No	L180349	
Calibration Date	17-08-2020	
Calibration Due Date	16-08-2021	
Traceability Certificate No	CC223920600022818 F (Validity- 16-08-2021)	
Calibration Done For	DISTRICT WOMEN HOSPITAL, PATHOLOGY LAB, PRATAPGARH	
Method of calibration	Calibration done using calibrated digital tachometer	
Results		
Instrument Reading (RPM)	Tachometer Reading (RPM)	% Variation
1000	1002	0.2
1500	1501	0.06
2000	2005	0.25
2500	2504	0.16
3000	3003	0.1

**Acceptance Criteria:** Less than  $\pm 5\%$

**Reference:** SMILE Johns Hopkins University Baltimore, MD USA  
Centrifuge Calibration Verification - Guidelines

  
**Calibration Done By:**  
Mr. Shashank Yadav  
Quality Specialist  
POCT Services

  
**Name & Signature:**  
Lab Incharge  
डिस्ट्रिक्ट महिला पथोलॉजी लैब  
प्रतापगढ़