

सिपेट : इंस्टीट्यूट ऑफ प्लास्टिक्स टेक्नोलॉजी (आई.पी.टी.)
CIPET : INSTITUTE OF PLASTICS TECHNOLOGY (IPT)
 (रसायन एवं पेट्रोसायन विभाग)
 (Department Of Chemicals & Petrochemicals)



CIPET सिपेट
 probe • perform • practice • Plastics

(रसायन एवं उर्वरक मंत्रालय, भारत सरकार)
 (Ministry Of Chemicals & Fertilizers, Govt. of India)
 बी-२७, अमौसी इण्डस्ट्रियल एरिया, लखनऊ-२२६ ००८
 B - 27, Amausi Industrial Area, Lucknow - 226 008



CC-2745

ग्राहक सेवा/Customer Service: 8756992882 फोन/Phone : +91-522 2436227
 ई-मेल/e-mail: cipetlko2@gmail.com, cipetcalibration1314@gmail.com
 website: www.cipet.gov.in

अंशांकन प्रमाण पत्र / Calibration Certificate

ULR No. : CC27452000001504F

CERTIFICATE NO. :	212(01)/TG/0320	SHEET 1 OF 1	
CALIBRATION OF :	Digital Thermometer	CALIBRATION DUE:	15.03.2021
CALIBRATED ON :	15.03.2020		

1. Calibrated for : M/S District Women Hospital
Moradabad
2. Service Request No. & date :
3. Identification of Device Under Calibration (DUC) :
 - a) Range : -50 to 300 °C
 - b) Least Count : 0.1 °C
 - c) Calibrated Range : 2 to 8 °C
 - d) Make/Model : MEXTECH / Multi thermometer
 - e) Serial No. : M18037797
 - f) Location : PL/DWHM/TM-02



4. Standard & Major equipment(s) used for calibration :

Standard(s)	Range/Leastcount	Limit of Accuracy/ Uncertainty in meas.	Traceability/ Validity
PRT Sensor with Digital Indicator & Low Temperature Bath	-40 to 600 °C 0.001 °C	(+/-) 0.34 °C	16(01)/TC/0519 12.05.2020

5. Environmental Condition(s) : Temperature : 23±2°C Relative Humidity: 50±10%
6. Calibration Procedure: CCC/CP-T/01

7. Calibration Result:

Sl. No.	Scale	DUC Value	Std.Value (Avg)	Correction
a.	°C	2	1.9842	-0.0158
b.	°C	5	4.9862	-0.0138
c.	°C	8	8.0038	0.0038

**Expanded Uncertainty of measurement = ± 0.37 °C
 with coverage factor k=2 at 95.45 % of confidence level.

8. Inference :

- (a) The correction is positive except when indicated otherwise, it has to be algebraically added to the DUC value in order to get actual value.
- (b) The calibration certificate shall not be reproduced except in full without written approval of laboratory ..
- (c) The above result related only to the instrument submitted in as received condition .
- (d) The above values are reported as such and may be taken accordingly whenever in use.
- (e) Label affixed on device

Checked by

Approved by
 Authorized Signatory

VIVEK KUMAR
 Sr. Technical Officer
 (Quality Manager)