

## LAB MONTHLY SUMMARY

Lab Name **CHENNAI DIAGNOSTIC CENTRE** Lab No **10389**  
 Month **August** Year **2020**  
 Constituent **Chemistry II**  
 Group

Date of Result Entered : 18/08/2020

Date of Report Published : 10/09/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Robonik ( Semi automatic / Fully Automatic )	362	112.00	10.66	11.94	112 mg/dl	0.00	1.26
2	UREA I	UREASE UV / GLDH	Robonik ( Semi automatic / Fully Automatic )	100	70.76	11.01	7.79	85 mg/dl	1.83	1.56
3	CREATININE I	JAFFE RATE-BLANKED &COMPENSATED/Alkaline Picrate	Robonik ( Semi automatic / Fully Automatic )	288	8.87	12.50	1.11	10 mg/dl	1.02	0.13
4	T.BILIRUBIN I	DIAZONIUM SALT ( Colorimetric) /JENDRASSIK	Robonik ( Semi automatic / Fully Automatic )	152	5.58	15.41	0.86	6.81 mg/dl	1.43	0.14
5	T-PROTEIN I	BIURET - colorimetric	Robonik ( Semi automatic / Fully Automatic )	151	5.54	10.98	0.61	4.9 g/dl	-1.05	0.10
6	ALBUMIN I	BCG - colorimetric	Robonik ( Semi automatic / Fully Automatic )	134	3.21	9.83	0.32	2.9 g/dl	-0.98	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Robonik ( Semi automatic / Fully Automatic )	137	6.00	11.16	0.67	5.8 mg/dl	-0.30	0.11
8	CHOLESTEROL I	CHOD-PAP	Robonik ( Semi automatic / Fully Automatic )	322	109.94	12.22	13.44	125 mg/dl	1.12	1.50
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Robonik ( Semi automatic / Fully Automatic )	173	259.24	14.86	38.52	350 mg/dl	2.36	5.86

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1


Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

