

# CALIBRATION CERTIFICATE

Identification of Instrument: Refrigerator Thermometer			
Make/Model No:	CF500	Serial no:	CF500170200708
Calibration Done on:	08-09-2020	Calibration Due Date:	09-09-2021
Reference Standard: Calibrated Thermometer			
Make/Model No:	Mextech/Multi Thermometer	Serial no:	M20883268
Calibration Done on:	15-03-2020	Calibration Due Date:	15-03-2021
Calibrated For: Pathology Lab, Pt. Kamla Pati Tripathi, District Female Hospital, Chandauli			
Certificate no: 213(01)/TG/0320			

Temp. of Ref Thermometer	2°C	4°C	6°C	8°C
S. No.	Temp.reading of Device	Temp.reading of Device	Temp.reading of Device	Temp.reading of Device
1	2.21	4.21	6.41	8.53
2	2.26	4.26	6.32	8.59
3	2.24	4.41	6.55	8.65
4	2.32	4.48	6.64	8.51
<b>Avg</b>	<b>2.2575</b>	<b>4.34</b>	<b>6.48</b>	<b>8.57</b>

**Calibration Result:**

SN.	Scale	Standard Value	Device Value (Avg)	Correction
1	Degree Celcius	2	2.25	0.25
2	Degree Celcius	4	4.34	0.34
3	Degree Celcius	6	6.48	0.48
4	Degree Celcius	8	6.57	0.57

Reference: SMILE Johns Hopkins University Baltimore, MD USA

*Shashank Yadav*  
 Calibration Done By  
 Mr. Shashank Yadav  
 POCT Services

*[Signature]*  
 Name & Signature  
 Lab Incharge

*[Signature]*  
 Pathologist  
 Dist. Hospital, Chandauli

REDMI NOTE 3 PRO MI DUAL CAMERA