


# CALIBRATION CERTIFICATE CENTRIFUGE


<b>Instrument Details</b>		
Instrument Name	Centrifuge	
Model	Remi R-8C	
Serial No	ZDJN-30058	
Calibration Date	08-09-2020	
Calibration Due Date	09-09-2021	
<b>Reference Standard Details</b>		
Reference Instrument Name	Tachometer	
Model No	Lutron/2234C	
Serial No	L180349	
Calibration Date	17-08-2020	
Calibration Due Date	16-08-2020	
Traceability Certificate No	CC223920600022818 F (Validity- 16-08-2021)	
Calibration Done For	Pathology Lab, Pt. Kamla Pati Tripathi, District Hospital, Chandauli	
Method of calibration	Calibration done using calibrated digital tachometer	
<b>Results</b>		
Instrument Reading (RPM)	Tachometer Reading (RPM)	% Variation
1500	1526	1.70
2000	2034	1.67
2500	2561	2.38
3000	3029	0.96
3500	3526	0.74
4000	4039	0.97
4500	4526	0.57

Acceptance Criteria: Less than  $\pm 5\%$

Reference: SMILE Johns Hopkins University Baltimore, MD USA  
Centrifuge Calibration Verification - Guidelines

  
 Calibration Done By:



  
 Dr. Shashank Jaiswal  
 Pathologist

Name & Signature:  
 Lab Incharge