

CALIBRATION CERTIFICATE

To,

Chennai Diagnostic Centre

Peravallur, Chennai-82

Instrument : AutoQuant 100 Amara

Machine Serial Number : AQ10150987

Issued On: 01-Feb-2020

Valid Upto : 31-Jan-2021

Appearance	No Imbalance Probe movement
Ambient temp(15-32)degree	24 Degree
Lamp Check	Ok
Threshold Value	35400(>25000)
Sample Dispersing Volume	In range (0 to 30ul)
Reagent Cooling Temperature	8 Degree(0 to 15 Degree)
Optical Path length	0.65
Valve, Motor & CRU	Working properly
Power Supply & Earthing	225volts& Grounding Less Than 5Volts

Conclusion: All Parameter are within range. The Machine is working properly (As per Meril terms)

Engineer Name : Vishnu G

Signature & Seal :



Meril Life Sciences India Pvt. Ltd | CIN : U33110GJ1999PTC060787

New no. 45, Old no. 23, 1st Floor, First Main Road, Shency Nagar (W), Chennai - 600 030 (TN)

T : +91 4277 4054 | F : +91 44 4277 4054 | E : askinfo@merilife.com | W : www.merilife.com

Registered Office: Bilakha House, Surver No: 135/139, Muktanand Marg, Chala, Vapi - 396 191 - Gujarat, India

Cardiovascular | Orthopedic | Diagnostics | Endo-Surgery | ENT