

## VIEW LAB MONTHLY SUMMARY

Lab Name	SHRI HANUMAN BALAJI CHARITABLE DIAGNOSTIC CENTER	Details About Robust Analysis
Lab No	11428	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2020	
Constituent Group	Chemistry I	

Click on the analyte to view Graphical Data

All Method Result

Print

Date of Result Entered : 10/08/2020

Date of Report Published : 10/09/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	245	107.92	5.88	6.35	98 mg/dl	-1.56	0.81
2	UREA	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	246	74.23	4.21	3.13	78 mg/dl	1.21	0.40
3	CREATININE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	247	9.91	4.34	0.43	9.8 mg/dl	-0.26	0.05
4	T.BILIRUBIN	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	240	5.79	5.76	0.33	5.4 mg/dl	-1.17	0.04
5	T-PROTEIN	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	249	4.97	4.89	0.24	4.9 g/dl	-0.29	0.03
6	ALBUMIN	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	252	2.74	3.97	0.11	2.8 g/dl	0.55	0.01
7	CALCIUM	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	244	11.11	3.10	0.34	11.2 mg/dl	0.26	0.04
8	URIC ACID	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	249	5.34	4.42	0.24	5.3 mg/dl	-0.17	0.03
9	CHOLESTEROL	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	231	100.70	5.15	5.19	98 mg/dl	-0.52	0.68
10	TRIGLYCERIDE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	214	343.89	4.06	13.96	330 mg/dl	-1.00	1.91
11	HDL CHO	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	217	20.55	7.77	1.60	20 mg/dl	-0.34	0.22
12	SODIUM	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	192	131.27	2.41	3.17	130 mmol/L	-0.40	0.46
13	POTASSIUM	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	190	4.53	3.07	0.14	4.5 mmol/L	-0.22	0.02
14	CHLORIDE	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	162	101.79	2.48	2.52	103 mmol/L	0.48	0.40
15	AST	Dry Chemistry with PLP (P-5-P)	JOHNSON & JOHNSON DRYCHEM SERIES	231	59.32	9.37	5.56	46 U/L	-2.40	0.73
16	ALT	Dry Chemistry	JOHNSON &	211	317.96	13.80	43.88	161 U/L	-3.58	6.04

		with PLP (P-5-P)	JOHNSON DRYCHEM SERIES							
17	ALP	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	235	147.04	9.51	13.98	148 U/L	0.07	1.82

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>