

246, Okhla Industrial Estate Phase-III,

New Delhi 110020, India

Tel : +91 (11) 4646 5000 / 4669 5001

Fax : +91 (11) 4646 5020 http : //www.horiba.com

CIN: U73100DL2006PTC153232

HIN/MED/2020-2021/100421 24th June 2020

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **PENTRA XL 80** bearing serial number: **811PXL8479** installed at **SHRI HANUMAN BALAJI CHARITABLE CENTRE, DELHI** as calibrated on 24th June 2020.

Calibrator : ABX MINOCAL

Lot No. : CX447

Expiry Date: 5th July 2020

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 23rd June 2021.



Shrish Dixit

(Head- Products & Marketing) For **Horiba India Pvt. Ltd.**