

## LAB MONTHLY SUMMARY

Lab Name **JEEVAN DIAGNOSTIC CENTRE**Lab No **13184**Month **June**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **03/07/2020**Date of Report Published : **08/08/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	845	80.95	11.31	9.16	79 mg/dl	-0.21	0.63
2	UREA I	UREASE UV / GLDH	Any Analyser	513	49.75	12.40	6.17	49.3 mg/dl	-0.07	0.54
3	CREATININE I	OTHERS ( Any Other Principles / Methods )	Any Analyser	117	3.46	12.50	0.43	3.68 mg/dl	0.51	0.08
4	T-PROTEIN I	BIURET - colorimetric	Any Analyser	663	4.87	11.36	0.55	4.96 g/dl	0.16	0.04
5	ALBUMIN I	BCG - colorimetric	Any Analyser	633	2.98	10.47	0.31	2.5 g/dl	-1.54	0.02
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	561	4.60	21.29	0.98	3.6 mg/dl	-1.02	0.08
7	CHOLESTEROL I	CHOD-PAP	Any Analyser	790	97.69	15.59	15.23	93.6 mg/dl	-0.27	1.08
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	710	87.69	20.45	17.94	86 mg/dl	-0.09	1.35

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*