

Register Office :
212, Sakar Complex,
Sanstha-Vasahat, Dandiabazar,
Raopura, Vadodara-390001, Guj.

E : info@shaktiscientific.in
Ph : +91 265 -2712565 / 2412756
M.: 87587 411 85 / 96386 39 393

CALIBRATION CERTIFICATE

CUSTOMER. Pulsepath Clinical Laboratory
1st Floor, 23, Aroma Complex Soma
Talav Road, Dabhoi Road, Vadodara

CERTIFICATE NO: 080

DESCRIPTION OF ITEM: Lab Centrifuge "Remi" **CERTIFICATE DATE:** 17/09/2020

MODEL NO: C854/6 **SR.NO:** ZEFN- 19115 **RANGE:** 3500RPM

DETAILS OF EQUIPMENT USED FOR CALIBRATION:

SR.NO	DESCRIPTION	CERTIFIED BY	CERTIFICAT NO.	VALIDITY
1	DIGITAL TECHOMETER	SADBHAV CALIBRATION	N/0311A/07/2020	27/07/2021

OBSERVATIONS



R.P.M.CALIBRATION:

POINT	INSTRUMENT TIMER DISPLAYED	CALIBRATED TECHO METER READING	ERROR OBSERVED	ACCEPTD ERROR	RESULT
1	1000RPM	980 RPM	- 20RPM	+/-100RPM	PASS
2	1500RPM	1495 RPM	- 05RPM	+/-100RPM	PASS
3	2000RPM	1986 RPM	- 14RPM	+/-100RPM	PASS
4.	3500RPM	3492RPM	-08RPM	+/-100RPM	PASS

1. Calibration done with Roto—6x15ml-----
2. Calibration Traceable to national Standards
3. Result is related only to the item calibrated
4. Maximum tolerance permissible RPM2%FSD



REMARKS: Overall deviation is within specified tolerance. Accepted
NEXT CALIBRATION DUE: 16/09/2021

CALIBRATED BY:


(SERVICE ENGR.)




APPROVED BY:


(DEPT. HEAD)


Corporate Office: LF-17, 4th Floor, Lotus Aura-1, Sama-Savli Road, Vemali, Vadodara-390008, Gujarat, INDIA.