

Register Office :
212, Sakar Complex,
Sanstha-Vasahat, Dandiabazar,
Raopura, Vadodara-390001, Guj.

E : info@shaktiscientific.in
Ph : +91 265 -2712565 / 2412756
M.: 87587 411 85 / 96386 39 393

CALIBRATION CERTIFICATE

CUSTOMER. Pulsepath Clinical Laboratory
1st Floor, 23, Aroma Complex Soma
Talav Road, Dabhoi Road, Vadodara

CERTIFICATE NO: 081

DESCRIPTION OF ITEM: Lab Centrifuge "Remi" **CERTIFICATE DATE:** 17/09/2020

MODEL NO: R8C

SR.NO: JFLC-11710

RANGE: 4000RPM

DETAILS OF EQUIPMENT USED FOR CALIBRATION:

SR.NO	DESCRIPTION	CERTIFIED BY	CERTIFICAT NO.	VALIDITY
1	DIGITAL TECHOMETER	SADBHAV CALIBRATION	N/0311A/07/2020	27/07/2021

OBSERVATIONS


R.P.M.CALIBRATION:

POINT	INSTRUMENT TIMER DISPLAYED	CALIBRATED TECHO-METER READING	ERROR OBSERVED	ACCEPTD ERROR	RESULT
1	1000RPM	994 RPM	- 06RPM	+/-100RPM	PASS
2	2000RPM	1993 RPM	- 07RPM	+/-100RPM	PASS
3	3000RPM	2988 RPM	- 12RPM	+/-100RPM	PASS
4.	4000RPM	3983RPM	-17RPM	+/-100RPM	PASS

1. Calibration done with Rotor Swing Out R81 16x15ml-----
2. Calibration Traceable to national Standards
3. Result is related only to the item calibrated
4. Maximum tolerance permissible RPM 2%FSD

REMARKS: Overall deviation is within specified tolerance. Accepted
NEXT CALIBRATION DUE: 16/09/2021

CALIBRATED BY:


(SERVICE ENGR.)




APPROVED BY:


(DEPT. HEAD)


Corporate Office: LF-17, 4th Floor, Lotus Aura-1, Sama-Savli Road, Vemali, Vadodara-390008, Gujarat, INDIA.