



NUTECH CALIBRATORS & ENGINEERS

(An NABL Accredited Calibration Laboratory in the field of Mechanical, Electro-Technical & Thermal)

Vide Certificate Nos.: CC-2271 (C-0421, C-0295 & C-0296)

Regd. Office & Lab. : 28/6 & 7, TOLLYGUNGE CIRCULAR ROAD, KOLKATA- 700 053

PHONE : (033) 2400 2182 (O), TELEFAX : 033 2400 2182

e-mail : nutech_call2006@rediffmail.com, nutech.calibration700053@gmail.com

Info@nutechcalibrators.com, Website : www.nutechcalibrators.com



NABL ACCREDITED
LABORATORY
CC-2271

Issued To:

M/s: Oncquest Laboratories Ltd. Kolkata
BF 2, Sushma Castle, Rajarhat Main Road,
Beside Arati Cinema Hall, P.O.: Jyangra,
Kolkata-700059, West Bengal.

Form No.- NCE/FM/03.

Certificate No.: NCE/CF-01854/20-21

ULR-CC227120000001099F

Date : 26/09/2020

CALIBRATION REPORT

CALIBRATED ON:
23/09/2020

NEXT DUE DATE :
22/09/2021

PAGE 1 OF 1

- 1) Reference Standard : TSC/18-19/9948-4
(Traceability No.& Date) Dt. 01/01/2019 and Valid up to 01/01/2021
- 2) Description Of Instrument : Centrifuge
a) Make : Remi b) Model : R-8C c) Range : 5250 rpm (Max) d) L/c : 10 rpm
- 3) Identification No : Sr. No.: ZDJN-27326
- 4) Receipt of Instrument : 23/09/2020
- 5) End User : Lab
- 6) Ambient condition during Calibration : a) Temp : 25°C ± 5°C b) Humidity : < 70 % RH
- 7) Basis of Calibration : Comparison Method
- 8) Standard/ Equipment used for Calibration : Digital Tachometer (NCE/DTM/01)
- 9) Measurement Uncertainty (of 95% C.L. at k = 2) : ± 0.15 % rpm (for the maximum reported range)
- 10) Laboratory Identification : Sticker is affixed for authentication of calibration.

CALIBRATION RESULTS

SI No.	DUC Set in rpm	Reading Observed through Ref. Std. in rpm
1.	2040	2046.2
2.	2760	2768.7
3.	3380	3389.5
4.	3820	3831.2
5.	4350	4364.6

REMARKS :The above DUC* has been calibrated over its range & the readings observed are tabulated above. The reference standard is traceable to National Standard. * DUC : Device Under Calibration

Calibrated By: S. Banik
S. Banik
Calibration Engineer

Approved By: S. GHOSH (QM) CUM TM
S. Ghosh
Authorised Signatory