

LAB MONTHLY SUMMARY

Lab Name **JEEVAN DIAGNOSTIC CENTRE**Lab No **13184**Month **February**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **19/02/2020**Date of Report Published : **06/03/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	942	91.40	10.04	9.17	86 mg/dl	-0.59	0.60
2	UREA I	UREASE UV / GLDH	Any Analyser	571	30.78	14.01	4.31	27.6 mg/dl	-0.74	0.36
3	CREATININE I	OTHERS (Any Other Principles / Methods)	Any Analyser	118	1.01	15.00	0.15	0.8 mg/dl	-1.39	0.03
4	T-PROTEIN I	BIURET - colorimetric	Any Analyser	737	4.75	12.62	0.60	4.3 g/dl	-0.75	0.04
5	ALBUMIN I	BCG - colorimetric	Any Analyser	702	2.95	11.91	0.35	2.2 g/dl	-2.14	0.03
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	679	4.74	19.40	0.92	4.5 mg/dl	-0.26	0.07
7	CHOLESTEROL I	CHOD-PAP	Any Analyser	786	96.71	12.41	12.00	104 mg/dl	0.61	0.86
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	796	130.14	13.94	18.14	138 mg/dl	0.43	1.29

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****